

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/30/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031316.D **Time Analyzed:** 14:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.318	9.227	9.427	746.630	712.500	4.8
2,4-D	8.440	8.346	8.546	737.630	705.000	4.6
2,4-DCAA	7.309	7.215	7.415	786.700	750.000	4.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/30/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031316.D **Time Analyzed:** 14:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.919	9.824	10.024	718.900	712.500	0.9
2,4-D	9.021	8.925	9.125	703.150	705.000	-0.3
2,4-DCAA	7.762	7.666	7.866	739.620	750.000	-1.4