

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/30/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031321.D **Time Analyzed:** 16:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.318	9.227	9.427	748.330	712.500	5.0
2,4-D	8.439	8.346	8.546	728.020	705.000	3.3
2,4-DCAA	7.309	7.215	7.415	782.490	750.000	4.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/29/2025 07/29/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/30/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031321.D **Time Analyzed:** 16:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.918	9.824	10.024	728.640	712.500	2.3
2,4-D	9.020	8.925	9.125	699.480	705.000	-0.8
2,4-DCAA	7.762	7.666	7.866	735.070	750.000	-2.0