

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS072125

Review By	Abdul	Review On	7/22/2025 7:57:36 AM
Supervise By	mohammad	Supervise On	7/23/2025 1:33:13 AM
SubDirectory	PS072125	HP Acquire Method	HP Processing Method ps072125 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031155.D	21 Jul 2025 14:14		AR\AJ	Ok
2	I.BLK	I.BLK	PS031156.D	21 Jul 2025 14:38		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS031157.D	21 Jul 2025 15:02		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS031158.D	21 Jul 2025 15:26		AR\AJ	Ok,M
5	HSTDICC750	HSTDICC750	PS031159.D	21 Jul 2025 15:51		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS031160.D	21 Jul 2025 16:15		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS031161.D	21 Jul 2025 16:39		AR\AJ	Ok
8	HSTDICV750	ICVPS072125	PS031162.D	21 Jul 2025 17:03		AR\AJ	Ok
9	I.BLK	I.BLK	PS031163.D	21 Jul 2025 17:27		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS031164.D	21 Jul 2025 17:51		AR\AJ	Ok
11	Q2529-10	TP-30	PS031165.D	21 Jul 2025 18:15	already analyzed	AR\AJ	Not Ok
12	Q2529-10MS	TP-30MS	PS031166.D	21 Jul 2025 18:40	some compound recovery fail ,already analyzed	AR\AJ	Not Ok
13	Q2529-10MSD	TP-30MSD	PS031167.D	21 Jul 2025 19:04	some compound recovery fail ,RPD fail,already analyzed	AR\AJ	Not Ok
14	PB168886BS	PB168886BS	PS031168.D	21 Jul 2025 19:28		AR\AJ	Ok
15	I.BLK	I.BLK	PS031169.D	21 Jul 2025 19:52		AR\AJ	Ok
16	HSTDCCC750	HSTDCCC750	PS031170.D	21 Jul 2025 20:16		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS072425

Review By	yogesh	Review On	7/25/2025 8:15:19 AM
Supervise By	mohammad	Supervise On	7/26/2025 3:49:45 AM
SubDirectory	PS072425	HP Acquire Method	HP Processing Method ps072125 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031220.D	24 Jul 2025 10:10		AR\AJ	Ok
2	I.BLK	I.BLK	PS031221.D	24 Jul 2025 10:35		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS031222.D	24 Jul 2025 12:03		AR\AJ	Ok,M
4	Q2638-11	OU4-TS-36-071725	PS031223.D	24 Jul 2025 13:44		AR\AJ	Ok
5	Q2638-11MS	OU4-TS-36-071725MS	PS031224.D	24 Jul 2025 14:08		AR\AJ	Ok,M
6	Q2638-11MSD	OU4-TS-36-071725MS	PS031225.D	24 Jul 2025 14:38		AR\AJ	Ok,M
7	Q2558-01MS	OU4-TS-Denali-070925	PS031226.D	24 Jul 2025 15:02		AR\AJ	Ok,M
8	Q2558-01MSD	OU4-TS-Denali-070925	PS031227.D	24 Jul 2025 15:26		AR\AJ	Ok,M
9	Q2638-13	OU4-TS-37-071725	PS031228.D	24 Jul 2025 15:50		AR\AJ	Ok,M
10	Q2638-05	OU4-TS-33-071725	PS031229.D	24 Jul 2025 16:14		AR\AJ	Ok
11	Q2638-07	OU4-TS-34-071725	PS031230.D	24 Jul 2025 16:39		AR\AJ	Ok
12	Q2641-02	P001-CONCRETE001-	PS031231.D	24 Jul 2025 17:03		AR\AJ	Ok
13	I.BLK	I.BLK	PS031232.D	24 Jul 2025 17:27		AR\AJ	Ok
14	HSTDCCC750	HSTDCCC750	PS031233.D	24 Jul 2025 18:39		AR\AJ	Ok,M
15	Q2641-02MS	P001-CONCRETE001-	PS031234.D	24 Jul 2025 19:04		AR\AJ	Ok,M
16	Q2641-02MSD	P001-CONCRETE001-	PS031235.D	24 Jul 2025 19:28		AR\AJ	Ok,M
17	PB169001BL	PB169001BL	PS031236.D	24 Jul 2025 19:52		AR\AJ	Ok
18	PB169001BS	PB169001BS	PS031237.D	24 Jul 2025 20:16		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS072425

Review By	yogesh	Review On	7/25/2025 8:15:19 AM
Supervise By	mohammad	Supervise On	7/26/2025 3:49:45 AM
SubDirectory	PS072425	HP Acquire Method	HP Processing Method ps072125 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC	PP24559		
Internal Standard/PEM			
ICV/I.BLK	PP24562		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PB168919TB	PB168919TB	PS031238.D	24 Jul 2025 20:40		AR\AJ	Ok
20	PB168926TB	PB168926TB	PS031239.D	24 Jul 2025 21:04		AR\AJ	Ok
21	PB168953TB	PB168953TB	PS031240.D	24 Jul 2025 21:29		AR\AJ	Ok
22	PB168969TB	PB168969TB	PS031241.D	24 Jul 2025 21:53		AR\AJ	Ok
23	I.BLK	I.BLK	PS031242.D	24 Jul 2025 22:17		AR\AJ	Ok
24	HSTDCCC750	HSTDCCC750	PS031243.D	24 Jul 2025 23:29		AR\AJ	Ok,M
25	Q2481-12	CC0627-AL	PS031244.D	24 Jul 2025 23:53	bad injection	AR\AJ	Not Ok
26	Q2481-15	CC0627-AOXL	PS031245.D	25 Jul 2025 00:17		AR\AJ	Ok,M
27	Q2667-01	C0AP2	PS031246.D	25 Jul 2025 00:41		AR\AJ	Ok
28	Q2667-02	C0AP3	PS031247.D	25 Jul 2025 01:05		AR\AJ	Ok
29	Q2481-19	CC0627-CLOXAL	PS031248.D	25 Jul 2025 01:29	f falg,	AR\AJ	Ok
30	Q2481-21	CC0627-SFBL	PS031249.D	25 Jul 2025 01:53		AR\AJ	Ok
31	Q2446-03	MR-BUR-LNG-13	PS031250.D	25 Jul 2025 02:18	bad injection, TYPO	AR\AJ	Not Ok
32	I.BLK	I.BLK	PS031251.D	25 Jul 2025 02:42		AR\AJ	Ok
33	HSTDCCC750	HSTDCCC750	PS031252.D	25 Jul 2025 03:54		AR\AJ	Ok,M
34	Q2481-13	CC0627-CLOXPL	PS031253.D	25 Jul 2025 04:18	Surrogate fail in 1st col.END CCC FAIL	AR\AJ	Ok
35	Q2481-14	CC0625-0XBL	PS031254.D	25 Jul 2025 04:43	Surrogate fail, END CCC FAIL, F flag	AR\AJ	Ok
36	Q2481-16	CC0625-NL	PS031255.D	25 Jul 2025 05:07	Surrogate fail, END CCC FAIL, F flag	AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS072425

Review By	yogesh	Review On	7/25/2025 8:15:19 AM
Supervise By	mohammad	Supervise On	7/26/2025 3:49:45 AM
SubDirectory	PS072425	HP Acquire Method	HP Processing Method ps072125 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

37	Q2481-17	CC0267-OXPL	PS031256.D	25 Jul 2025 05:31	Surrogate fail, END CCC FAIL, F flag	AR\AJ	Ok
38	Q2481-18	CC0627-OXL	PS031257.D	25 Jul 2025 05:55	bad injection, surrogate not detected	AR\AJ	Not Ok
39	Q2481-20	CC0627-BL	PS031258.D	25 Jul 2025 06:19	Surrogate fail 2nd col, END CCC FAIL, F flag	AR\AJ	Ok
40	I.BLK	I.BLK	PS031259.D	25 Jul 2025 06:43	Surrogate low in both column, bad injection	AR\AJ	Ok
41	HSTDCCC750	HSTDCCC750	PS031260.D	25 Jul 2025 07:07	most of compounds are fail	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS072825

Review By	Abdul	Review On	7/30/2025 8:54:35 AM
Supervise By	mohammad	Supervise On	7/31/2025 1:17:07 AM
SubDirectory	PS072825	HP Acquire Method	HP Processing Method pd072125 8081

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	I.BLK	I.BLK	PS031262.D	28 Jul 2025 09:34		AR\AJ	Ok
2	HSTDCCC750	HSTDCCC750	PS031263.D	28 Jul 2025 11:35		AR\AJ	Ok,M
3	Q2481-12	CC0627-AL	PS031264.D	28 Jul 2025 14:31	End CCC missing, only for fax	AR\AJ	ReRun
4	Q2481-13	CC0627-CLOXPL	PS031265.D	28 Jul 2025 14:55	End CCC missing	AR\AJ	Not Ok
5	Q2481-14	CC0625-OXBL	PS031266.D	28 Jul 2025 15:43	End CCC missing	AR\AJ	Not Ok
6	Q2481-16	CC0625-NL	PS031267.D	28 Jul 2025 16:32	End CCC missing	AR\AJ	Not Ok
7	Q2481-17	CC0267-OXPL	PS031268.D	28 Jul 2025 17:20	End CCC missing	AR\AJ	Not Ok
8	Q2481-18	CC0627-OXL	PS031269.D	28 Jul 2025 18:08	Surrogate fail,End CCC missing, f flag only for fax	AR\AJ	ReRun
9	Q2481-20	CC0627-BL	PS031270.D	28 Jul 2025 18:57	End CCC missing	AR\AJ	Not Ok
10	Q2481-19	CC0627-CLOXAL	PS031271.D	28 Jul 2025 07:45 pm	End CCC missing	AR\AJ	Not Ok
11	Q2481-21	CC0627-SFBL	PS031272.D	28 Jul 2025 20:33	End CCC missing	AR\AJ	Not Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS072925

Review By	Abdul	Review On	7/30/2025 8:35:13 AM
Supervise By	mohammad	Supervise On	7/31/2025 1:17:44 AM
SubDirectory	PS072925	HP Acquire Method	HP Processing Method ps072925 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031273.D	29 Jul 2025 15:42		AR\AJ	Ok
2	I.BLK	I.BLK	PS031274.D	29 Jul 2025 16:06		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS031275.D	29 Jul 2025 16:30		AR\AJ	Ok,M
4	HSTDICC500	HSTDICC500	PS031276.D	29 Jul 2025 16:54		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS031277.D	29 Jul 2025 17:18		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS031278.D	29 Jul 2025 17:42		AR\AJ	Ok,M
7	HSTDICC1500	HSTDICC1500	PS031279.D	29 Jul 2025 18:07		AR\AJ	Ok,M
8	HSTDICV750	ICVPS072925	PS031280.D	29 Jul 2025 18:31		AR\AJ	Ok
9	I.BLK	I.BLK	PS031281.D	29 Jul 2025 18:55		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS031282.D	29 Jul 2025 19:19		AR\AJ	Ok,M
11	Q2668-01	TP-2	PS031283.D	29 Jul 2025 19:43		AR\AJ	Ok
12	Q2668-05	TP-3	PS031284.D	29 Jul 2025 20:07		AR\AJ	Ok
13	Q2668-09	TP-1	PS031285.D	29 Jul 2025 20:32		AR\AJ	Ok
14	PB169024BL	PB169024BL	PS031286.D	29 Jul 2025 20:56		AR\AJ	Ok
15	PB169024BS	PB169024BS	PS031287.D	29 Jul 2025 21:20		AR\AJ	Ok,M
16	Q2668-09MS	TP-1MS	PS031288.D	29 Jul 2025 21:44	Some compound recovery fail	AR\AJ	Ok,M
17	Q2668-09MSD	TP-1MSD	PS031289.D	29 Jul 2025 22:08	Some compound recovery fail	AR\AJ	Ok,M
18	Q2691-01	295	PS031290.D	29 Jul 2025 22:32	surrogate fail in both column	AR\AJ	Not Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS072925

Review By	Abdul	Review On	7/30/2025 8:35:13 AM
Supervise By	mohammad	Supervise On	7/31/2025 1:17:44 AM
SubDirectory	PS072925	HP Acquire Method	HP Processing Method ps072925 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

19	Q2691-07	299	PS031291.D	29 Jul 2025 22:57	Surrogate fail in 1st col	AR\AJ	ReRun
20	I.BLK	I.BLK	PS031292.D	29 Jul 2025 23:21		AR\AJ	Ok
21	HSTDCCC750	HSTDCCC750	PS031293.D	29 Jul 2025 23:45		AR\AJ	Ok,M
22	PB169037BL	PB169037BL	PS031294.D	30 Jul 2025 00:09		AR\AJ	Ok
23	PB169037BS	PB169037BS	PS031295.D	30 Jul 2025 00:33		AR\AJ	Ok
24	PB168986TB	PB168986TB	PS031296.D	30 Jul 2025 00:57		AR\AJ	Ok
25	Q2681-01	NYPA-POUCH-SPENT	PS031297.D	30 Jul 2025 01:20		AR\AJ	Ok
26	Q2681-01MS	NYPA-POUCH-SPENT	PS031298.D	30 Jul 2025 01:44	Recovery fail	AR\AJ	Ok,M
27	Q2681-01MSD	NYPA-POUCH-SPENT	PS031299.D	30 Jul 2025 02:09	Recovery fail	AR\AJ	Ok,M
28	I.BLK	I.BLK	PS031300.D	30 Jul 2025 02:33		AR\AJ	Ok
29	HSTDCCC750	HSTDCCC750	PS031301.D	30 Jul 2025 02:57		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS073025

Review By	yogesh	Review On	7/31/2025 8:24:42 AM
Supervise By	mohammad	Supervise On	7/31/2025 8:26:32 AM
SubDirectory	PS073025	HP Acquire Method	HP Processing Method ps072925 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC Internal Standard/PEM	PP24559
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031302.D	30 Jul 2025 08:15		AR\AJ	Ok
2	I.BLK	I.BLK	PS031303.D	30 Jul 2025 08:40		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS031304.D	30 Jul 2025 09:04		AR\AJ	Ok,M
4	Q2703-01	TP-4	PS031305.D	30 Jul 2025 09:28		AR\AJ	Ok
5	Q2706-01	RT-5417	PS031306.D	30 Jul 2025 09:52		AR\AJ	Ok,M
6	Q2706-03	ETGI-361	PS031307.D	30 Jul 2025 10:16		AR\AJ	Ok,M
7	I.BLK	I.BLK	PS031308.D	30 Jul 2025 10:41		AR\AJ	Ok
8	HSTDCCC750	HSTDCCC750	PS031309.D	30 Jul 2025 11:05		AR\AJ	Ok,M
9	Q2709-01	7211935-295	PS031310.D	30 Jul 2025 11:35		AR\AJ	Ok,M
10	Q2709-01MS	7211935-295MS	PS031311.D	30 Jul 2025 11:59	some compound recovery fail	AR\AJ	Ok,M
11	Q2709-01MSD	7211935-295MSD	PS031312.D	30 Jul 2025 12:23	some compound recovery fail , RPD fail	AR\AJ	Ok,M
12	PB169051BL	PB169051BL	PS031313.D	30 Jul 2025 12:49		AR\AJ	Ok
13	PB169051BS	PB169051BS	PS031314.D	30 Jul 2025 13:13		AR\AJ	Ok,M
14	I.BLK	I.BLK	PS031315.D	30 Jul 2025 13:37		AR\AJ	Ok
15	HSTDCCC750	HSTDCCC750	PS031316.D	30 Jul 2025 14:01		AR\AJ	Ok,M
16	Q2481-12	CC0627-AL	PS031317.D	30 Jul 2025 14:25		AR\AJ	Ok
17	Q2481-18	CC0627-OXL	PS031318.D	30 Jul 2025 14:49	Surrogate high in 1st column	AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS073025

Review By	yogesh	Review On	7/31/2025 8:24:42 AM
Supervise By	mohammad	Supervise On	7/31/2025 8:26:32 AM
SubDirectory	PS073025	HP Acquire Method	HP Processing Method ps072925 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC	PP24559		
Internal Standard/PEM			
ICV/I.BLK	PP24562		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q2691-07RE	299RE	PS031319.D	30 Jul 2025 15:38	Surrogate fail in 1st col	AR\AJ	Confirms
19	I.BLK	I.BLK	PS031320.D	30 Jul 2025 16:02		AR\AJ	Ok
20	HSTDCCC750	HSTDCCC750	PS031321.D	30 Jul 2025 16:26		AR\AJ	Ok,M
21	Q2691-01	295	PS031322.D	30 Jul 2025 16:51		AR\AJ	Ok
22	Q2639-01	OU4-TS-38-071725	PS031323.D	30 Jul 2025 17:15		AR\AJ	Ok,M
23	Q2639-03	OU4-TS-39-071725	PS031324.D	30 Jul 2025 17:39		AR\AJ	Ok,M
24	Q2639-05RE	OU4-TS-40-071725RE	PS031325.D	30 Jul 2025 18:03	Surrogate low in both column	AR\AJ	Confirms
25	Q2639-09RE	OU4-TS-42-071725RE	PS031326.D	30 Jul 2025 18:27	Surrogate low in 1st column	AR\AJ	Confirms
26	Q2639-11RE	OU4-TS-43-071725RE	PS031327.D	30 Jul 2025 18:51	Surrogate low in 1st column	AR\AJ	Confirms
27	Q2639-13RE	OU4-TS-44-071725RE	PS031328.D	30 Jul 2025 19:15	Surrogate low in 1st column	AR\AJ	Confirms
28	I.BLK	I.BLK	PS031329.D	30 Jul 2025 19:40		AR\AJ	Ok
29	HSTDCCC750	HSTDCCC750	PS031330.D	30 Jul 2025 22:05		AR\AJ	Ok,M

M : Manual Integration