

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/29/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089668.D **Time Analyzed:** 14:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.081	8.971	9.171	56.770	50.000	13.5
Endrin	6.582	6.472	6.672	54.490	50.000	9.0
gamma-BHC (Lindane)	4.338	4.228	4.428	53.490	50.000	7.0
Heptachlor	4.937	4.827	5.027	48.240	50.000	-3.5
Heptachlor epoxide	5.699	5.588	5.788	53.780	50.000	7.6
Methoxychlor	7.501	7.392	7.592	50.280	50.000	0.6
Tetrachloro-m-xylene	3.557	3.448	3.648	50.420	50.000	0.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/29/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089668.D **Time Analyzed:** 14:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.070	7.970	8.170	46.190	50.000	-7.6
Endrin	5.787	5.687	5.887	45.910	50.000	-8.2
gamma-BHC (Lindane)	3.728	3.627	3.827	47.390	50.000	-5.2
Heptachlor	4.081	3.980	4.180	44.990	50.000	-10.0
Heptachlor epoxide	4.871	4.770	4.970	46.820	50.000	-6.4
Methoxychlor	6.752	6.651	6.851	41.450	50.000	-17.1
Tetrachloro-m-xylene	2.878	2.779	2.979	46.680	50.000	-6.6