

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/29/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089675.D **Time Analyzed:** 18:41

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.079	8.971	9.171	48.600	50.000	-2.8
Endrin	6.580	6.472	6.672	53.670	50.000	7.3
gamma-BHC (Lindane)	4.336	4.228	4.428	53.510	50.000	7.0
Heptachlor	4.935	4.827	5.027	46.050	50.000	-7.9
Heptachlor epoxide	5.697	5.588	5.788	53.610	50.000	7.2
Methoxychlor	7.499	7.392	7.592	49.550	50.000	-0.9
Tetrachloro-m-xylene	3.556	3.448	3.648	49.960	50.000	-0.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/29/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089675.D **Time Analyzed:** 18:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.070	7.970	8.170	40.880	50.000	-18.2
Endrin	5.787	5.687	5.887	45.650	50.000	-8.7
gamma-BHC (Lindane)	3.727	3.627	3.827	47.480	50.000	-5.0
Heptachlor	4.080	3.980	4.180	44.760	50.000	-10.5
Heptachlor epoxide	4.870	4.770	4.970	47.240	50.000	-5.5
Methoxychlor	6.752	6.651	6.851	41.420	50.000	-17.2
Tetrachloro-m-xylene	2.878	2.779	2.979	46.880	50.000	-6.2