

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/24/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096568.D **Time Analyzed:** 18:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.016	8.920	9.120	57.800	50.000	15.6
Endrin	6.543	6.445	6.645	57.090	50.000	14.2
gamma-BHC (Lindane)	4.312	4.213	4.413	56.850	50.000	13.7
Heptachlor	4.903	4.806	5.006	57.150	50.000	14.3
Heptachlor epoxide	5.663	5.565	5.765	57.210	50.000	14.4
Methoxychlor	7.461	7.364	7.564	59.290	50.000	18.6
Tetrachloro-m-xylene	3.536	3.438	3.638	56.140	50.000	12.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/24/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096568.D **Time Analyzed:** 18:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.993	7.896	8.096	38.640	50.000	-22.7
Endrin	5.710	5.612	5.812	54.160	50.000	8.3
gamma-BHC (Lindane)	3.666	3.567	3.767	58.700	50.000	17.4
Heptachlor	4.015	3.917	4.117	56.670	50.000	13.3
Heptachlor epoxide	4.800	4.702	4.902	55.590	50.000	11.2
Methoxychlor	6.682	6.584	6.784	59.020	50.000	18.0
Tetrachloro-m-xylene	2.829	2.729	2.929	58.000	50.000	16.0