

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/24/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096577.D **Time Analyzed:** 23:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.015	8.920	9.120	48.250	50.000	-3.5
Endrin	6.541	6.445	6.645	48.910	50.000	-2.2
gamma-BHC (Lindane)	4.311	4.213	4.413	53.040	50.000	6.1
Heptachlor	4.901	4.806	5.006	43.900	50.000	-12.2
Heptachlor epoxide	5.660	5.565	5.765	56.820	50.000	13.6
Methoxychlor	7.462	7.364	7.564	47.730	50.000	-4.5
Tetrachloro-m-xylene	3.535	3.438	3.638	49.130	50.000	-1.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2481
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/24/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096577.D **Time Analyzed:** 23:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.992	7.896	8.096	43.920	50.000	-12.2
Endrin	5.709	5.612	5.812	41.280	50.000	-17.4
gamma-BHC (Lindane)	3.665	3.567	3.767	53.260	50.000	6.5
Heptachlor	4.014	3.917	4.117	47.510	50.000	-5.0
Heptachlor epoxide	4.799	4.702	4.902	51.760	50.000	3.5
Methoxychlor	6.681	6.584	6.784	40.150	50.000	-19.7
Tetrachloro-m-xylene	2.828	2.729	2.929	52.110	50.000	4.2