

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF071725

Review By	Rahul	Review On	7/18/2025 11:02:11 AM
Supervise By	Jagrut	Supervise On	7/18/2025 1:30:54 PM
SubDirectory	BF071725	HP Acquire Method	BNA_F
		HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143138.D	17 Jul 2025 09:58		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143139.D	17 Jul 2025 10:28	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BF143140.D	17 Jul 2025 11:04		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BF143141.D	17 Jul 2025 11:34		RC/JU	Ok,M
5	SSTDICC010	SSTDICC010	BF143142.D	17 Jul 2025 12:04		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BF143143.D	17 Jul 2025 12:34		RC/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BF143144.D	17 Jul 2025 13:03	Compound #32,54,65 Kept on LR	RC/JU	Ok,M
8	SSTDICC050	SSTDICC050	BF143145.D	17 Jul 2025 13:33		RC/JU	Ok,M
9	SSTDICC060	SSTDICC060	BF143146.D	17 Jul 2025 14:04		RC/JU	Ok,M
10	SSTDICC080	SSTDICC080	BF143147.D	17 Jul 2025 14:34	Compound#77 removed from 80 ppm	RC/JU	Ok,M
11	SSTDICV040	ICVBF071725	BF143148.D	17 Jul 2025 15:06		RC/JU	Ok,M
12	PB168813BL	PB168813BL	BF143149.D	17 Jul 2025 15:36		RC/JU	Ok
13	Q2126-03	MDL-SOIL-03-QT2-202	BF143150.D	17 Jul 2025 16:06	MDL-SOIL 4 ppm	RC/JU	Ok,M
14	Q2126-03	MDL-SOIL-03-QT2-202	BF143151.D	17 Jul 2025 16:36	MDL-SOIL 8 ppm	RC/JU	Ok,M
15	Q2126-09	MDL-WATER-03-QT2-2	BF143152.D	17 Jul 2025 17:07	MDL-WATER 4 ppm	RC/JU	Ok,M
16	Q2126-09	MDL-WATER-03-QT2-2	BF143153.D	17 Jul 2025 17:37	MDL-WATER 8 ppm	RC/JU	Ok,M
17	PB168816BL	PB168816BL	BF143154.D	17 Jul 2025 18:07		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF071725

Review By	Rahul	Review On	7/18/2025 11:02:11 AM		
Supervise By	Jagrut	Supervise On	7/18/2025 1:30:54 PM		
SubDirectory	BF071725	HP Acquire Method	BNA_F	HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	PB167904BL	PB167904BL	BF143155.D	17 Jul 2025 18:37		RC/JU	Ok
19	PB167904BS	PB167904BS	BF143156.D	17 Jul 2025 19:07		RC/JU	Ok,M
20	SSTDCCC040	SSTDCCC040EC	BF143157.D	17 Jul 2025 19:37		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF072425

Review By	Rahul	Review On	7/25/2025 8:49:54 AM
Supervise By	Jagrut	Supervise On	7/25/2025 11:05:01 AM
SubDirectory	BF072425	HP Acquire Method	BNA_F
		HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13168,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143222.D	24 Jul 2025 09:03		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143223.D	24 Jul 2025 10:00		RC/JU	Ok,M
3	PB168981BL	PB168981BL	BF143224.D	24 Jul 2025 10:29		RC/JU	Ok
4	PB168981BS	PB168981BS	BF143225.D	24 Jul 2025 10:58		RC/JU	Ok,M
5	PB168983BL	PB168983BL	BF143226.D	24 Jul 2025 11:28		RC/JU	Ok
6	PB168983BS	PB168983BS	BF143227.D	24 Jul 2025 11:57		RC/JU	Ok,M
7	Q2668-08	TP-3	BF143228.D	24 Jul 2025 12:33		RC/JU	Ok
8	Q2668-12	TP-1	BF143229.D	24 Jul 2025 13:02		RC/JU	Ok
9	Q2481-21	CC0627-SFBL	BF143230.D	24 Jul 2025 13:32	Surrogate Fail	RC/JU	Ok
10	Q2668-01	TP-2	BF143231.D	24 Jul 2025 14:01		RC/JU	Ok
11	Q2668-04	TP-2	BF143232.D	24 Jul 2025 14:30		RC/JU	Ok
12	Q2668-04MS	TP-2MS	BF143233.D	24 Jul 2025 14:59		RC/JU	Ok,M
13	Q2668-04MSD	TP-2MSD	BF143234.D	24 Jul 2025 15:29		RC/JU	Ok,M
14	Q2668-09	TP-1	BF143235.D	24 Jul 2025 15:58		RC/JU	Ok
15	Q2668-09MS	TP-1MS	BF143236.D	24 Jul 2025 16:28		RC/JU	Ok
16	Q2668-09MSD	TP-1MSD	BF143237.D	24 Jul 2025 16:57		RC/JU	Ok
17	Q2676-01	HD-01-07222025	BF143238.D	24 Jul 2025 17:26		RC/JU	Ok,M
18	Q2674-01	TR-04-072225	BF143239.D	24 Jul 2025 17:56		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF072425

Review By	Rahul	Review On	7/25/2025 8:49:54 AM		
Supervise By	Jagrut	Supervise On	7/25/2025 11:05:01 AM		
SubDirectory	BF072425	HP Acquire Method	BNA_F	HP Processing Method	bf071725
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S13168,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2675-01	SU-03-07222025	BF143240.D	24 Jul 2025 18:25		RC/JU	Ok,M
20	Q2668-05	TP-3	BF143241.D	24 Jul 2025 18:55		RC/JU	Ok
21	Q2677-03	OK-03-07222025	BF143242.D	24 Jul 2025 19:24		RC/JU	Ok,M
22	Q2677-01	OK-02-07222025	BF143243.D	24 Jul 2025 19:53		RC/JU	Ok,M
23	Q2676-03	HD-02-07222025	BF143244.D	24 Jul 2025 20:22		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP072125

Review By	Rahul	Review On	7/22/2025 8:59:18 AM
Supervise By	Jagrut	Supervise On	7/22/2025 9:10:39 AM
SubDirectory	BP072125	HP Acquire Method	BNA_P
		HP Processing Method	bp072125

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025194.D	21 Jul 2025 13:02		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025195.D	21 Jul 2025 13:44	A Fresh Calibration is required.	CG/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BP025196.D	21 Jul 2025 14:25		CG/JU	Ok
4	SSTDICC005	SSTDICC005	BP025197.D	21 Jul 2025 15:06		CG/JU	Ok,M
5	SSTDICC010	SSTDICC010	BP025198.D	21 Jul 2025 15:47		CG/JU	Ok
6	SSTDICC020	SSTDICC020	BP025199.D	21 Jul 2025 16:29		CG/JU	Ok
7	SSTDICCC040	SSTDICCC040	BP025200.D	21 Jul 2025 17:10		CG/JU	Ok
8	SSTDICC050	SSTDICC050	BP025201.D	21 Jul 2025 17:51		CG/JU	Ok,M
9	SSTDICC060	SSTDICC060	BP025202.D	21 Jul 2025 18:32		CG/JU	Ok
10	SSTDICC080	SSTDICC080	BP025203.D	21 Jul 2025 19:14	Compound#79 removed from 80 PPM.	CG/JU	Ok,M
11	SSTDICV040	ICVBP072125	BP025204.D	21 Jul 2025 19:55		CG/JU	Ok,M
12	PB168854BL	PB168854BL	BP025205.D	21 Jul 2025 21:58		CG/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP072425

Review By	Rahul	Review On	7/25/2025 8:50:44 AM
Supervise By	Jagrut	Supervise On	7/25/2025 10:17:07 AM
SubDirectory	BP072425	HP Acquire Method	BNA_P
		HP Processing Method	bp072125

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13168,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025234.D	24 Jul 2025 09:13		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025235.D	24 Jul 2025 09:54		CG/JU	Ok,M
3	PB168969TB	PB168969TB	BP025236.D	24 Jul 2025 10:36		CG/JU	Ok
4	PB168953TB	PB168953TB	BP025237.D	24 Jul 2025 11:17		CG/JU	Ok
5	Q2481-12	CC0627-AL	BP025238.D	24 Jul 2025 12:01		CG/JU	Ok
6	Q2481-15	CC0627-AOXL	BP025239.D	24 Jul 2025 12:43	Poor recoveries of Internal standard & surrogates	CG/JU	Ok
7	Q2481-16	CC0625-NL	BP025240.D	24 Jul 2025 13:24		CG/JU	Ok
8	Q2481-14	CC0625-OXBL	BP025241.D	24 Jul 2025 14:05	Internal Standard Fail	CG/JU	Ok
9	Q2481-17	CC0267-OXPL	BP025242.D	24 Jul 2025 14:46	Poor recoveries of Internal standard	CG/JU	Ok
10	Q2481-20	CC0627-BL	BP025243.D	24 Jul 2025 15:27	Surrogate Fail	CG/JU	Ok
11	Q2481-19	CC0627-CLOXAL	BP025244.D	24 Jul 2025 16:09	Poor recoveries of Internal standard & surrogates	CG/JU	Ok
12	Q2481-13	CC0627-CLOXPL	BP025245.D	24 Jul 2025 16:50	Poor recoveries of Internal standard & surrogates	CG/JU	Ok,M
13	Q2481-18	CC0627-OXL	BP025246.D	24 Jul 2025 17:31	Poor recoveries of Internal standard & surrogates	CG/JU	Ok

M : Manual Integration