

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ENVI60</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2481</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>07/24/2025 09:44</u>
Lab File ID:	<u>VX047105.D</u>	Init. Calib. Date(s):	<u>07/21/2025 07/21/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:47 11:32</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.706	0.789		11.76	20
1,1-Dichloroethene	0.628	0.640		1.91	20
2-Butanone	0.363	0.345		-4.96	20
Carbon Tetrachloride	0.562	0.551		-1.96	20
Chloroform	1.236	1.218		-1.46	20
Benzene	1.523	1.505		-1.18	20
1,2-Dichloroethane	0.536	0.537		0.19	20
Trichloroethene	0.391	0.374		-4.35	20
Tetrachloroethene	0.360	0.338		-6.11	20
Chlorobenzene	1.171	1.122	0.3	-4.18	20
1,2-Dichloroethane-d4	0.623	0.644		3.37	20
Dibromofluoromethane	0.309	0.335		8.41	20
Toluene-d8	1.000	1.006		0.6	20
4-Bromofluorobenzene	0.431	0.444		3.02	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.