

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX072125

Review By	John Carlone	Review On	7/22/2025 8:51:02 AM
Supervise By	Maresh Dadoda	Supervise On	7/22/2025 1:41:20 PM
SubDirectory	VX072125	HP Acquire Method	HP Processing Method 82X072125W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134843
Initial Calibration Stds	VP134844,VP134845,VP134846,VP134847,VP134848,VP134849
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP134850
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047054.D	21 Jul 2025 08:53		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX047055.D	21 Jul 2025 09:47		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX047056.D	21 Jul 2025 10:08	Comp. #11 is on Linear Regression	JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX047057.D	21 Jul 2025 10:29		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX047058.D	21 Jul 2025 10:50		JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX047059.D	21 Jul 2025 11:11		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX047060.D	21 Jul 2025 11:32		JC/MD	Ok,M
8	IBLK	IBLK	VX047061.D	21 Jul 2025 11:52		JC/MD	Ok
9	VSTDICV050	ICVVX072125	VX047062.D	21 Jul 2025 13:17		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX072325

Review By	John Carlone	Review On	7/24/2025 8:47:51 AM
Supervise By	Maresh Dadoda	Supervise On	7/24/2025 3:06:37 PM
SubDirectory	VX072325	HP Acquire Method	HP Processing Method 82X072125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134888
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134889,VP134890

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047076.D	23 Jul 2025 09:00		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047077.D	23 Jul 2025 09:30	pH#Lot#V12668	JC/MD	Ok,M
3	VX0723MBL01	VX0723MBL01	VX047078.D	23 Jul 2025 09:58		JC/MD	Ok
4	VX0723WBL01	VX0723WBL01	VX047079.D	23 Jul 2025 10:45		JC/MD	Ok
5	VX0723WBS01	VX0723WBS01	VX047080.D	23 Jul 2025 11:13		JC/MD	Ok,M
6	VX0723MBS01	VX0723MBS01	VX047081.D	23 Jul 2025 11:41		JC/MD	Ok,M
7	PB168956TB	PB168956TB	VX047082.D	23 Jul 2025 12:02		JC/MD	Ok,M
8	Q2668-04	TP-2	VX047083.D	23 Jul 2025 12:22	vial A pH#5.0	JC/MD	Ok,M
9	Q2668-08	TP-3	VX047084.D	23 Jul 2025 12:43	vial A pH#5.0	JC/MD	Ok
10	Q2668-12	TP-1	VX047085.D	23 Jul 2025 13:04	vial A pH#5.0	JC/MD	Ok
11	Q2672-01	AUD-25-0124	VX047086.D	23 Jul 2025 13:25	vial A pH#5.0	JC/MD	Ok,M
12	Q2672-03	AUD-25-0128	VX047087.D	23 Jul 2025 13:46	vial A pH#5.0	JC/MD	Ok
13	VX0723WBSD01	VX0723WBSD01	VX047088.D	23 Jul 2025 14:07		JC/MD	Ok,M
14	Q2660-03	MOO-25-0218	VX047089.D	23 Jul 2025 14:28	Run @ Lower Dilution	JC/MD	Not Ok
15	IBLK	IBLK	VX047090.D	23 Jul 2025 14:49		JC/MD	Ok
16	Q2660-03	MOO-25-0218	VX047091.D	23 Jul 2025 15:10	cloth material	JC/MD	Ok,M
17	Q2481-15	CC0627-AOXL	VX047092.D	23 Jul 2025 15:31	Need Straight Run	JC/MD	Not Ok
18	Q2481-21	CC0627-SFBL	VX047093.D	23 Jul 2025 15:52	Surrogate fail;Need Straight Run	JC/MD	Not Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX072325

Review By	John Carlone	Review On	7/24/2025 8:47:51 AM
Supervise By	Maresh Dadoda	Supervise On	7/24/2025 3:06:37 PM
SubDirectory	VX072325	HP Acquire Method	HP Processing Method 82X072125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134888		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134889,VP134890		

19	Q2481-16	CC0625-NL	VX047094.D	23 Jul 2025 16:13	Need Straight Run	JC/MD	Not Ok
20	Q2481-20	CC0627-BL	VX047095.D	23 Jul 2025 16:34	Need Straight Run	JC/MD	Not Ok
21	Q2481-14	CC0625-OXBL	VX047096.D	23 Jul 2025 16:54	Need Straight Run	JC/MD	Not Ok
22	Q2481-12	CC0627-AL	VX047097.D	23 Jul 2025 17:15	Need Straight Run	JC/MD	Not Ok
23	PB168966TB	PB168966TB	VX047098.D	23 Jul 2025 17:36		JC/MD	Ok,M
24	Q2481-19	CC0627-CLOXAL	VX047099.D	23 Jul 2025 17:57	Need Straight Run	JC/MD	Not Ok
25	Q2481-13	CC0627-CLOXPL	VX047100.D	23 Jul 2025 18:18	Need Straight Run	JC/MD	Not Ok
26	Q2481-17	CC0267-OXPL	VX047101.D	23 Jul 2025 18:39	Need Straight Run	JC/MD	Not Ok
27	Q2481-18	CC0627-OXL	VX047102.D	23 Jul 2025 19:00	vial A pH<2	JC/MD	Ok,M
28	VSTDCCC050	VSTDCCC050EC	VX047103.D	23 Jul 2025 19:21		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX072425

Review By	John Carlone	Review On	7/25/2025 8:51:46 AM
Supervise By	Maresh Dadoda	Supervise On	7/25/2025 4:38:01 PM
SubDirectory	VX072425	HP Acquire Method	HP Processing Method 82X072125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134905
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134906,VP134907

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047104.D	24 Jul 2025 09:12		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047105.D	24 Jul 2025 09:44	pH#Lot#V12668	JC/MD	Ok,M
3	VX0724MBL01	VX0724MBL01	VX047106.D	24 Jul 2025 10:13		JC/MD	Ok
4	VX0724WBL01	VX0724WBL01	VX047107.D	24 Jul 2025 10:34		JC/MD	Ok
5	VX0724WBS01	VX0724WBS01	VX047108.D	24 Jul 2025 10:55		JC/MD	Ok,M
6	VX0724WBSD01	VX0724WBSD01	VX047109.D	24 Jul 2025 11:21		JC/MD	Ok,M
7	Q2682-01	TOWER-1	VX047110.D	24 Jul 2025 11:42	vial A pH<2	JC/MD	Ok
8	Q2682-02	TOWER-2	VX047111.D	24 Jul 2025 12:03	vial A pH<2 Surrogate Fail	JC/MD	ReRun
9	IBLK	IBLK	VX047112.D	24 Jul 2025 12:24		JC/MD	Ok
10	Q2682-02	TOWER-2	VX047113.D	24 Jul 2025 12:45	vial B pH<2	JC/MD	Ok
11	Q2481-14	CC0625-OXBL	VX047114.D	24 Jul 2025 13:06	vial A pH#6.0 Surrogate Fail	JC/MD	Ok
12	Q2481-13	CC0627-CLOXPL	VX047115.D	24 Jul 2025 13:26	vial A pH<2	JC/MD	Ok,M
13	IBLK	IBLK	VX047116.D	24 Jul 2025 13:47		JC/MD	Ok
14	Q2481-12	CC0627-AL	VX047117.D	24 Jul 2025 14:08	vial A pH<2	JC/MD	Ok
15	Q2481-15	CC0627-AOXL	VX047118.D	24 Jul 2025 14:29	vial A pH<2 Surrogate Fail	JC/MD	Ok
16	Q2481-16	CC0625-NL	VX047119.D	24 Jul 2025 14:50	vial A pH#6.0	JC/MD	Ok
17	Q2481-21	CC0627-SFBL	VX047120.D	24 Jul 2025 15:12	vial A pH#6.0 Surrogate Fail; Internal Standard Fail	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX072425

Review By	John Carlone	Review On	7/25/2025 8:51:46 AM
Supervise By	Maresh Dadoda	Supervise On	7/25/2025 4:38:01 PM
SubDirectory	VX072425	HP Acquire Method	HP Processing Method 82X072125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134905		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134906,VP134907		

18	Q2481-19	CC0627-CLOXAL	VX047121.D	24 Jul 2025 15:33	vial A pH<2	JC/MD	Ok
19	Q2481-20	CC0627-BL	VX047122.D	24 Jul 2025 15:54	vial A pH#6.0 Surrogate Fail;Internal Standard Fail	JC/MD	Ok
20	Q2481-17	CC0267-OXPL	VX047123.D	24 Jul 2025 16:15	vial A pH<2	JC/MD	Ok
21	IBLK	IBLK	VX047124.D	24 Jul 2025 16:37		JC/MD	Ok
22	IBLK	IBLK	VX047125.D	24 Jul 2025 16:58		JC/MD	Ok
23	VSTDCCC050	VSTDCCC050EC	VX047126.D	24 Jul 2025 17:19		JC/MD	Ok,M

M : Manual Integration