

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

|                 |                                              |                           |                          |
|-----------------|----------------------------------------------|---------------------------|--------------------------|
| <b>SDG No.:</b> | <u><b>Q2481</b></u>                          | <b>Analytical Method:</b> | <u><b>SW8260D</b></u>    |
| <b>Client:</b>  | <u><b>Environmental Restoration, LLC</b></u> | <b>Datafile :</b>         | <u><b>VX047080.D</b></u> |

| Lab Sample ID      | Parameter            | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits<br>High | RPD |
|--------------------|----------------------|-------|--------|------|-----|-----|------|-----|----------------|-----|
| <b>VX0723WBS01</b> | Vinyl chloride       | 20    | 19.1   | ug/L | 96  |     |      | 65  | 117            |     |
|                    | 1,1-Dichloroethene   | 20    | 18.7   | ug/L | 94  |     |      | 74  | 110            |     |
|                    | 2-Butanone           | 100   | 88.9   | ug/L | 89  |     |      | 65  | 122            |     |
|                    | Carbon Tetrachloride | 20    | 18.1   | ug/L | 91  |     |      | 77  | 113            |     |
|                    | Chloroform           | 20    | 18.9   | ug/L | 95  |     |      | 79  | 113            |     |
|                    | Benzene              | 20    | 18.7   | ug/L | 94  |     |      | 82  | 109            |     |
|                    | 1,2-Dichloroethane   | 20    | 19.2   | ug/L | 96  |     |      | 80  | 115            |     |
|                    | Trichloroethene      | 20    | 18.0   | ug/L | 90  |     |      | 77  | 113            |     |
|                    | Tetrachloroethene    | 20    | 18.1   | ug/L | 91  |     |      | 67  | 123            |     |
|                    | Chlorobenzene        | 20    | 18.8   | ug/L | 94  |     |      | 82  | 109            |     |