

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENVI60
 Lab Code: ACE SDG NO.: Q2481
 Lab File ID: VX047077.D Date Analyzed: 07/23/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:30
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	346372	5.56	582956	6.76	499261	10.05
UPPER LIMIT	692744	6.056	1165910	7.263	998522	10.549
LOWER LIMIT	173186	5.056	291478	6.263	249631	9.549
EPA SAMPLE NO.						
CC0627-OXL	416777	5.56	772965	6.77	519074	10.06
VX0723WBL01	495715	5.56	888951	6.76	838806	10.06
VX0723WBS01	361496	5.56	612680	6.76	542704	10.06
VX0723WBSD01	310959	5.56	528421	6.77	477316	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.