

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>ENVI60</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2481</u>
Lab File ID:	<u>VX047105.D</u>	Date Analyzed:	<u>07/24/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:44</u>
GC Column:	<u>DB-624UI</u> ID: <u>0.18</u> (mm)	Heated Purge:	(Y/N) <u>N</u>

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	299639	5.56	500897	6.76	440622	10.06
UPPER LIMIT	599278	6.056	1001790	7.263	881244	10.555
LOWER LIMIT	149820	5.056	250449	6.263	220311	9.555
EPA SAMPLE NO.						
CC0627-AL	405182	5.56	744375	6.77	710158	10.06
CC0627-CLOXPL	306378	5.62	578026	6.81	434831	10.08
CC0625-0XBL	452175	5.56	826688	6.77	795791	10.06
CC0627-AOXL	401016	5.56	749846	6.77	708181	10.06
CC0625-NL	396009	5.56	749209	6.77	689935	10.06
CC0267-0XPL	354820	5.58	677494	6.78	521532	10.06
CC0627-CLOXAL	414773	5.56	745400	6.77	702427	10.06
CC0627-BL	503250	5.56	912608	6.77	857469	10.06
CC0627-SFBL	488832	5.56	892867	6.77	845132	10.06
VX0724WBL01	463916	5.56	839242	6.77	802389	10.06
VX0724WBS01	305867	5.57	519601	6.77	466090	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.