

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENVI60
 Lab Code: ACE SDG NO.: Q2481
 Lab File ID: VX047077.D Date Analyzed: 07/23/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:30
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	238699	12.018				
UPPER LIMIT	477398	12.518				
LOWER LIMIT	119350	11.518				
EPA SAMPLE NO.						
CC0627-OXL	323235	12.02				
VX0723WBL01	406646	12.02				
VX0723WBS01	268316	12.02				
VX0723WBSD01	242466	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.