

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>ENVI60</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2481</u>
Lab File ID:	<u>VX047105.D</u>	Date Analyzed:	<u>07/24/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:44</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	207085	12.018				
UPPER LIMIT	414170	12.518				
LOWER LIMIT	103543	11.518				
EPA SAMPLE NO.						
CC0627-AL	363979	12.02				
CC0627-CLOXPL	264294	12.03				
CC0625-OXBL	405857	12.02				
CC0627-AOXL	355368	12.02				
CC0625-NL	372334	12.02				
CC0267-OXPL	319546	12.02				
CC0627-CLOXAL	362645	12.02				
CC0627-BL	440518 *	12.02				
CC0627-SFBL	418999 *	12.02				
VX0724WBL01	396966	12.02				
VX0724WBS01	230557	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.