


LB136369

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.7541e-004

C= -6.4789e-002

Rho= 0.9998275

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%/.
0.0	0.000	-0.059	-0.059	34	0.000	34					
0.2	0.200	0.194	-0.006	1476	0.0 %	1476					
2.5	2.500	2.517	0.017	14720	0.0 %	14720					
5.0	5.000	5.062	0.062	29228	0.0 %	29228					
7.5	7.500	7.594	0.094	43661	0.0 %	43661					
10.0	10.000	9.892	-0.108	56761	0.0 %	56761					

LB136369 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Jul 2025 10:53:01

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Std Conc	Method	Date
S	0.0 - 1	SO	-	34	PPB	0.0000	7471B	03 Jul 2025 10:55:45
S	0.2 - 1	SV 2	-	1476	PPB	0.2000	7471B	03 Jul 2025 10:58:02
S	2.5 - 1	SV 5	-	14720	PPB	2.5000	7471B	03 Jul 2025 11:00:20
S	5.0 - 1	SV 5	-	29228	PPB	5.0000	7471B	03 Jul 2025 11:02:37
S	7.5 - 1	SV 5	-	43661	PPB	7.5000	7471B	03 Jul 2025 11:04:57
S	10.0 - 1	SV	-	56761	PPB	10.0000	7471B	03 Jul 2025 11:10:06
U	ICV16 - 1	ICV16	4.2594	24652	PPB	-7471B		03 Jul 2025 11:13:05
U	ICB16 - 1	ICB16	-0.0639	5	PPB	-7471B		03 Jul 2025 11:17:58
U	CCV51 - 1	CCV51	5.1438	29694	PPB	-7471B		03 Jul 2025 11:20:12
U	CCB51 - 1	CCB51	-0.0560	50	PPB	-7471B		03 Jul 2025 11:22:28
U	CRA - 1	CRA	0.1908	1457	PPB	-7471B		03 Jul 2025 11:24:48
U	HighStd - 1	HighStd	10.0825	57849	PPB	-7471B		03 Jul 2025 11:27:03
U	ChkStd - 1	ChkStd	6.6720	38406	PPB	-7471B		03 Jul 2025 11:29:18
U	PB168719BS - 1	LCSS	4.1985	24305	PPB	-7471B		03 Jul 2025 11:34:14
U	Q2473-01 - 1	PIT#1	2.0554	12087	PPB	-7471B		03 Jul 2025 11:36:31
U	Q2473-02 - 1	PIT#2	2.6425	15434	PPB	-7471B		03 Jul 2025 11:38:52
U	Q2473-03 - 1	PIT#3	1.9568	11525	PPB	-7471B		03 Jul 2025 11:41:11
U	Q2473-04 - 1	PIT#4	2.5590	14958	PPB	-7471B		03 Jul 2025 11:43:31
U	Q2475-01 - 1	SOIL-PILE	0.4653	3022	PPB	-7471B		03 Jul 2025 11:45:48
U	CCV52 - 1	CCV52	5.2254	30159	PPB	-7471B		03 Jul 2025 11:48:06
U	CCB52 - 1	CCB52	-0.0318	188	PPB	-7471B		03 Jul 2025 11:50:26
U	Q2478-01 - 1	WC-1	0.5614	3570	PPB	-7471B		03 Jul 2025 11:52:51
U	Q2484-01 - 1	TP-58	0.1692	1334	PPB	-7471B		03 Jul 2025 11:55:06
U	Q2484-02 - 1	TP-57	0.9247	5641	PPB	-7471B		03 Jul 2025 11:57:24
U	Q2484-03 - 1	TP-64	0.2825	1980	PPB	-7471B		03 Jul 2025 11:59:39
U	Q2484-04 - 1	TP-107	0.7117	4427	PPB	-7471B		03 Jul 2025 12:01:57
U	Q2484-05 - 1	TP-1006	0.3837	2557	PPB	-7471B		03 Jul 2025 12:04:13
U	Q2484-06 - 1	TP-104	0.5191	3329	PPB	-7471B		03 Jul 2025 12:06:32
U	Q2484-06DUP - 1	TP-104DUP	0.4588	2985	PPB	-7471B		03 Jul 2025 12:08:50
U	Q2484-06MS - 1	TP-104MS	4.9130	28378	PPB	-7471B		03 Jul 2025 12:11:08
U	Q2484-06MSD - 1	TP-104MSD	4.9786	28752	PPB	-7471B		03 Jul 2025 12:13:25
U	CCV53 - 1	CCV53	5.0223	29001	PPB	-7471B		03 Jul 2025 12:15:43
U	CCB53 - 1	CCB53	-0.0499	85	PPB	-7471B		03 Jul 2025 12:18:04
U	PB168719BL - 1	PBS	-0.0236	235	PPB	-7471B		03 Jul 2025 12:20:25
U	Q2484-06LX5 - 1		0.0568	693	PPB	-7471B		03 Jul 2025 12:26:25
U	Q2484-06A - 1		4.6741	27016	PPB	-7471B		03 Jul 2025 12:28:40
U	CCV54 - 1	CCV54	5.0389	29096	PPB	-7471B		03 Jul 2025 12:30:55
U	CCB54 - 1	CCB54	-0.0686	-22	PPB	-7471B		03 Jul 2025 12:33:16