

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

ProjectID: South River WM Replacement

Lab Code: ACE

SDG No.: Q2484

Calibration Sequence : FF061325		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	128762116	128762	FF015964.D
500	64295722	128591	FF015965.D
200	27169561	135848	FF015966.D
100	13526499	135265	FF015967.D
50	7306219	146124	FF015968.D
AVG RF : 134918		% RSD : 5.299	AVG RT : 15.0944