

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2484
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/07/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030932.D **Time Analyzed:** 21:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.652	9.568	9.768	762.650	712.500	7.0
2,4,5-TP(Silvex)	9.358	9.275	9.475	709.800	712.500	-0.4
2,4-D	8.471	8.386	8.586	683.020	705.000	-3.1
2,4-DB	10.229	10.147	10.347	797.070	712.500	11.9
2,4-DCAA	7.336	7.249	7.449	697.140	750.000	-7.0
DICAMBA	7.526	7.439	7.639	684.890	705.000	-2.9
DICHLORPROP	8.239	8.153	8.353	641.940	705.000	-8.9
Dinoseb	11.444	11.366	11.566	693.310	705.000	-1.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2484
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/07/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030932.D **Time Analyzed:** 21:36

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.355	10.259	10.459	688.820	712.500	-3.3
2,4,5-TP(Silvex)	9.928	9.833	10.033	680.850	712.500	-4.4
2,4-D	9.024	8.927	9.127	639.250	705.000	-9.3
2,4-DB	10.922	10.827	11.027	673.080	712.500	-5.5
2,4-DCAA	7.766	7.670	7.870	694.830	750.000	-7.4
DICAMBA	7.969	7.872	8.072	671.350	705.000	-4.8
DICHLORPROP	8.688	8.591	8.791	636.650	705.000	-9.7
Dinoseb	11.307	11.211	11.411	651.850	705.000	-7.5