

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2484
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/08/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030941.D **Time Analyzed:** 01:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.651	9.568	9.768	776.510	712.500	9.0
2,4,5-TP(Silvex)	9.357	9.275	9.475	716.040	712.500	0.5
2,4-D	8.470	8.386	8.586	695.710	705.000	-1.3
2,4-DB	10.227	10.147	10.347	806.190	712.500	13.1
2,4-DCAA	7.336	7.249	7.449	703.920	750.000	-6.1
DICAMBA	7.526	7.439	7.639	686.270	705.000	-2.7
DICHLORPROP	8.238	8.153	8.353	648.310	705.000	-8.0
Dinoseb	11.444	11.366	11.566	685.440	705.000	-2.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2484
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/08/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030941.D **Time Analyzed:** 01:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.354	10.259	10.459	691.170	712.500	-3.0
2,4,5-TP(Silvex)	9.928	9.833	10.033	682.700	712.500	-4.2
2,4-D	9.023	8.927	9.127	645.590	705.000	-8.4
2,4-DB	10.922	10.827	11.027	665.750	712.500	-6.6
2,4-DCAA	7.766	7.670	7.870	692.550	750.000	-7.7
DICAMBA	7.969	7.872	8.072	673.190	705.000	-4.5
DICHLORPROP	8.688	8.591	8.791	637.580	705.000	-9.6
Dinoseb	11.306	11.211	11.411	636.320	705.000	-9.7