

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2484
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/09/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030967.D **Time Analyzed:** 14:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.653	9.568	9.768	709.820	712.500	-0.4
2,4,5-TP(Silvex)	9.359	9.275	9.475	668.070	712.500	-6.2
2,4-D	8.471	8.386	8.586	643.940	705.000	-8.7
2,4-DB	10.231	10.147	10.347	759.090	712.500	6.5
2,4-DCAA	7.337	7.249	7.449	668.070	750.000	-10.9
DICAMBA	7.526	7.439	7.639	648.820	705.000	-8.0
DICHLORPROP	8.240	8.153	8.353	607.260	705.000	-13.9
Dinoseb	11.447	11.366	11.566	644.130	705.000	-8.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2484
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/09/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030967.D **Time Analyzed:** 14:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.361	10.259	10.459	635.110	712.500	-10.9
2,4,5-TP(Silvex)	9.934	9.833	10.033	629.760	712.500	-11.6
2,4-D	9.028	8.927	9.127	626.500	705.000	-11.1
2,4-DB	10.929	10.827	11.027	617.890	712.500	-13.3
2,4-DCAA	7.769	7.670	7.870	636.310	750.000	-15.2
DICAMBA	7.972	7.872	8.072	613.280	705.000	-13.0
DICHLORPROP	8.692	8.591	8.791	588.860	705.000	-16.5
Dinoseb	11.313	11.211	11.411	588.950	705.000	-16.5