

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CAMP02  
**Lab Code:** ACE **SDG NO.:** Q2484  
**GC Column:** RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

**Client Sample No.:** CCAL07 **Date Analyzed:** 07/10/2025  
**Lab Sample No.:** HSTDCCC750 **Data File :** PS030980.D **Time Analyzed:** 09:48

| COMPOUND         | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|----------------------|--------|--------------------|-------------------|------|
| 2,4,5-T          | 9.652  | 9.568                | 9.768  | 767.030            | 712.500           | 7.7  |
| 2,4,5-TP(Silvex) | 9.356  | 9.275                | 9.475  | 706.790            | 712.500           | -0.8 |
| 2,4-D            | 8.470  | 8.386                | 8.586  | 689.810            | 705.000           | -2.2 |
| 2,4-DB           | 10.229 | 10.147               | 10.347 | 835.670            | 712.500           | 17.3 |
| 2,4-DCAA         | 7.336  | 7.249                | 7.449  | 701.710            | 750.000           | -6.4 |
| DICAMBA          | 7.525  | 7.439                | 7.639  | 677.750            | 705.000           | -3.9 |
| DICHLORPROP      | 8.238  | 8.153                | 8.353  | 639.810            | 705.000           | -9.2 |
| Dinoseb          | 11.445 | 11.366               | 11.566 | 678.890            | 705.000           | -3.7 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CAMP02  
**Lab Code:** ACE **SDG NO.:** Q2484  
**GC Column:** RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

**Client Sample No.:** CCAL07 **Date Analyzed:** 07/10/2025  
**Lab Sample No.:** HSTDCCC750 **Data File :** PS030980.D **Time Analyzed:** 09:48

| COMPOUND         | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|------------------|--------|----------------------|--------|--------------------|-------------------|-------|
| 2,4,5-T          | 10.360 | 10.259               | 10.459 | 657.200            | 712.500           | -7.8  |
| 2,4,5-TP(Silvex) | 9.934  | 9.833                | 10.033 | 652.700            | 712.500           | -8.4  |
| 2,4-D            | 9.027  | 8.927                | 9.127  | 643.630            | 705.000           | -8.7  |
| 2,4-DB           | 10.928 | 10.827               | 11.027 | 639.290            | 712.500           | -10.3 |
| 2,4-DCAA         | 7.769  | 7.670                | 7.870  | 654.250            | 750.000           | -12.8 |
| DICAMBA          | 7.972  | 7.872                | 8.072  | 635.160            | 705.000           | -9.9  |
| DICHLORPROP      | 8.691  | 8.591                | 8.791  | 616.740            | 705.000           | -12.5 |
| Dinoseb          | 11.313 | 11.211               | 11.411 | 618.480            | 705.000           | -12.3 |