

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2484
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL08 **Date Analyzed:** 07/10/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030990.D **Time Analyzed:** 16:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.650	9.568	9.768	782.090	712.500	9.8
2,4,5-TP(Silvex)	9.355	9.275	9.475	720.380	712.500	1.1
2,4-D	8.469	8.386	8.586	702.820	705.000	-0.3
2,4-DB	10.227	10.147	10.347	848.520	712.500	19.1
2,4-DCAA	7.334	7.249	7.449	718.940	750.000	-4.1
DICAMBA	7.524	7.439	7.639	694.160	705.000	-1.5
DICHLORPROP	8.237	8.153	8.353	649.970	705.000	-7.8
Dinoseb	11.443	11.366	11.566	705.470	705.000	0.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2484
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL08 **Date Analyzed:** 07/10/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030990.D **Time Analyzed:** 16:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.359	10.259	10.459	678.370	712.500	-4.8
2,4,5-TP(Silvex)	9.932	9.833	10.033	671.710	712.500	-5.7
2,4-D	9.027	8.927	9.127	622.580	705.000	-11.7
2,4-DB	10.927	10.827	11.027	653.050	712.500	-8.3
2,4-DCAA	7.768	7.670	7.870	672.370	750.000	-10.4
DICAMBA	7.971	7.872	8.072	654.990	705.000	-7.1
DICHLORPROP	8.691	8.591	8.791	633.040	705.000	-10.2
Dinoseb	11.312	11.211	11.411	643.930	705.000	-8.7