

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
 Data File : PS030742.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2025 12:37
 Operator : AR\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/19/2025
 Supervised By :mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 13:32:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 18 13:31:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.349	7.770	3449.9E6	993.0E6	1007.482m	978.598
Target Compounds							
1) T	Dalapon	2.711	2.710	5150.3E6	2361.2E6	890.676	891.083
2) T	3,5-DICHL...	6.508	6.716	4617.8E6	1397.4E6	912.175	911.477
3) T	4-Nitroph...	7.149	7.302	1346.0E6	1506.7E6	887.706	886.926
5) T	DICAMBA	7.540	7.973	13729.6E6	5900.9E6	922.935	921.488
6) T	MCP	7.723	8.073	915.4E6	207.0E6	90.761	91.801
7) T	MCPA	7.873	8.323	1089.6E6	295.3E6	89.789	91.284
8) T	DICHLORPROP	8.254	8.692	3084.8E6	1328.0E6	916.366	923.462
9) T	2,4-D	8.486	9.027	2961.7E6	1432.3E6	916.252	915.890
10) T	Pentachlo...	8.799	9.553	44933.5E6	34544.9E6	970.273	964.535
11) T	2,4,5-TP ...	9.375	9.933	17920.7E6	12790.2E6	930.890	933.380
12) T	2,4,5-T	9.669	10.359	15378.5E6	12286.7E6	928.339	932.902
13) T	2,4-DB	10.248	10.927	2257.2E6	1053.3E6	919.502	932.595
14) T	DINOSEB	11.468	11.313	12449.4E6	9605.3E6	918.844	921.811
15) T	Picloram	11.275	12.421	15192.8E6	21456.6E6	916.121	928.345
16) T	DCPA	11.763	12.357	23325.9E6	19939.4E6	946.473	946.862

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061825\
Data File : PS030742.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2025 12:37
Operator : AR\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/19/2025
Supervised By : mohammad ahmed 06/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 13:32:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061825.M
Quant Title : 8080.M
QLast Update : Wed Jun 18 13:31:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

