

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2484</u>	Analytical Method:	<u>8151A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PS030949.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits		RPD
			Result	Result			Qual	RPD	Qual	Low	High		
Lab Sample ID:	Q2458-01MS (Column 1)	Client Sample ID:		TP-76MS									
	DICAMBA	183.4	0	86.9	ug/Kg	47					10	112	
	DICHLORPROP	183.4	0	95.6	ug/Kg	52					10	113	
	2,4-D	183.4	0	138	ug/Kg	75					10	144	
	2,4,5-TP(Silvex)	183.4	0	95.3	ug/Kg	52					10	114	
	2,4,5-T	183.4	0	99.8	ug/Kg	54					10	115	
	2,4-DB	183.4	0	97.1	ug/Kg	53					10	140	
	Dinoseb	183.4	0	0	ug/Kg	0	*				10	118	
Lab Sample ID:	Q2458-01MS (Column 2)	Client Sample ID:		TP-76MS									
	DICAMBA	183.4	0	84.8	ug/Kg	46					10	112	
	DICHLORPROP	183.4	0	103	ug/Kg	56					10	113	
	2,4-D	183.4	0	94.4	ug/Kg	51					10	144	
	2,4,5-TP(Silvex)	183.4	0	97.3	ug/Kg	53					10	114	
	2,4,5-T	183.4	0	91.0	ug/Kg	50					10	115	
	2,4-DB	183.4	0	85.1	ug/Kg	46					10	140	
	Dinoseb	183.4	0	0	ug/Kg	0	*				10	118	