

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>CAMP02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2484</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):	<u>07/01/2025</u> <u>07/01/2025</u>
		Calibration Times:	<u>14:04</u> <u>21:30</u>

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PP073413.D</u>	RT 750 = <u>PP073414.D</u>
RT 500 = <u>PP073415.D</u>	RT 250 = <u>PP073416.D</u>	RT 050 = <u>PP073417.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	5.65	5.64	5.64	5.64	5.64	5.64	5.54	5.74
Aroclor-1016-2	(2)	5.67	5.66	5.67	5.66	5.66	5.66	5.56	5.76
Aroclor-1016-3	(3)	5.73	5.72	5.73	5.72	5.73	5.73	5.63	5.83
Aroclor-1016-4	(4)	5.83	5.82	5.83	5.82	5.82	5.82	5.72	5.92
Aroclor-1016-5	(5)	6.12	6.11	6.12	6.11	6.12	6.12	6.02	6.22
Aroclor-1260-1	(1)	7.24	7.23	7.23	7.23	7.23	7.23	7.13	7.33
Aroclor-1260-2	(2)	7.49	7.48	7.49	7.48	7.49	7.49	7.39	7.59
Aroclor-1260-3	(3)	7.85	7.84	7.85	7.84	7.84	7.85	7.75	7.95
Aroclor-1260-4	(4)	8.07	8.07	8.07	8.07	8.07	8.07	7.97	8.17
Aroclor-1260-5	(5)	8.39	8.38	8.39	8.38	8.39	8.39	8.29	8.49
Decachlorobiphenyl		10.19	10.18	10.19	10.18	10.18	10.18	10.08	10.28
Tetrachloro-m-xylene		4.50	4.49	4.49	4.49	4.49	4.49	4.39	4.59
Aroclor-1242-1	(1)	5.64	5.64	5.65	5.64	5.64	5.64	5.54	5.74
Aroclor-1242-2	(2)	5.67	5.66	5.67	5.66	5.67	5.66	5.56	5.76
Aroclor-1242-3	(3)	5.73	5.72	5.73	5.72	5.73	5.73	5.63	5.83
Aroclor-1242-4	(4)	5.82	5.82	5.83	5.82	5.83	5.82	5.72	5.92
Aroclor-1242-5	(5)	6.55	6.55	6.56	6.55	6.55	6.55	6.45	6.65
Decachlorobiphenyl		10.19	10.18	10.19	10.18	10.19	10.18	10.08	10.28
Tetrachloro-m-xylene		4.49	4.49	4.49	4.49	4.49	4.49	4.39	4.59
Aroclor-1248-1	(1)	5.64	5.64	5.64	5.64	5.64	5.64	5.54	5.74
Aroclor-1248-2	(2)	5.91	5.91	5.91	5.92	5.91	5.91	5.81	6.01
Aroclor-1248-3	(3)	6.12	6.12	6.11	6.12	6.12	6.12	6.02	6.22
Aroclor-1248-4	(4)	6.51	6.51	6.51	6.52	6.51	6.51	6.41	6.61
Aroclor-1248-5	(5)	6.55	6.55	6.55	6.56	6.55	6.55	6.45	6.65
Decachlorobiphenyl		10.18	10.18	10.18	10.19	10.18	10.18	10.08	10.28
Tetrachloro-m-xylene		4.49	4.49	4.49	4.49	4.49	4.49	4.39	4.59
Aroclor-1254-1	(1)	6.49	6.49	6.49	6.49	6.49	6.49	6.39	6.59
Aroclor-1254-2	(2)	6.70	6.71	6.71	6.70	6.70	6.71	6.61	6.81
Aroclor-1254-3	(3)	7.07	7.07	7.07	7.07	7.07	7.07	6.97	7.17
Aroclor-1254-4	(4)	7.35	7.35	7.35	7.35	7.35	7.35	7.25	7.45
Aroclor-1254-5	(5)	7.77	7.77	7.77	7.77	7.77	7.77	7.67	7.87
Decachlorobiphenyl		10.18	10.18	10.18	10.18	10.18	10.18	10.08	10.28
Tetrachloro-m-xylene		4.49	4.49	4.49	4.49	4.49	4.49	4.39	4.59
Aroclor-1268-1	(1)	8.70	8.69	8.70	8.69	8.69	8.69	8.59	8.79
Aroclor-1268-2	(2)	8.79	8.79	8.79	8.79	8.79	8.79	8.69	8.89
Aroclor-1268-3	(3)	9.02	9.01	9.02	9.02	9.01	9.02	8.92	9.12
Aroclor-1268-4	(4)	9.44	9.43	9.44	9.43	9.43	9.43	9.33	9.53
Aroclor-1268-5	(5)	9.85	9.84	9.85	9.84	9.84	9.84	9.74	9.94

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: **CAMP02**

Lab Code: ACE

SDG NO.: Q2484

Instrument ID: ECD_P

Calibration Date(s): **07/01/2025** **07/01/2025**

Calibration Times: 14:04 21:30

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 =	<u>PP073413.D</u>	RT 750 =	<u>PP073414.D</u>
RT 500 =	PP073415.D	RT 250 =	PP073416.D	RT 050 = PP073417.D

[illegible]

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

[illegible]