

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2484
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073448.D **Time Analyzed:** 10:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.651	5.544	5.744	569.830	500.000	14.0
Aroclor-1016-2	5.673	5.566	5.766	563.490	500.000	12.7
Aroclor-1016-3	5.735	5.629	5.829	542.290	500.000	8.5
Aroclor-1016-4	5.832	5.726	5.926	557.220	500.000	11.4
Aroclor-1016-5	6.124	6.018	6.218	561.610	500.000	12.3
Aroclor-1260-1	7.241	7.134	7.334	539.260	500.000	7.9
Aroclor-1260-2	7.494	7.389	7.589	562.220	500.000	12.4
Aroclor-1260-3	7.853	7.746	7.946	543.520	500.000	8.7
Aroclor-1260-4	8.076	7.971	8.171	528.530	500.000	5.7
Aroclor-1260-5	8.394	8.289	8.489	541.310	500.000	8.3
Decachlorobiphenyl	10.193	10.086	10.286	54.850	50.000	9.7
Tetrachloro-m-xylene	4.498	4.393	4.593	56.360	50.000	12.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2484
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073448.D **Time Analyzed:** 10:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.864	4.764	4.964	480.890	500.000	-3.8
Aroclor-1016-2	4.881	4.782	4.982	483.670	500.000	-3.3
Aroclor-1016-3	5.058	4.958	5.158	495.110	500.000	-1.0
Aroclor-1016-4	5.100	5.000	5.200	491.380	500.000	-1.7
Aroclor-1016-5	5.314	5.214	5.414	500.010	500.000	0.0
Aroclor-1260-1	6.345	6.244	6.444	485.850	500.000	-2.8
Aroclor-1260-2	6.533	6.433	6.633	443.500	500.000	-11.3
Aroclor-1260-3	6.685	6.585	6.785	495.020	500.000	-1.0
Aroclor-1260-4	7.155	7.054	7.254	490.060	500.000	-2.0
Aroclor-1260-5	7.397	7.297	7.497	480.740	500.000	-3.9
Decachlorobiphenyl	8.790	8.690	8.890	51.990	50.000	4.0
Tetrachloro-m-xylene	3.783	3.685	3.885	46.450	50.000	-7.1