

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2484
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073478.D **Time Analyzed:** 23:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.640	5.544	5.744	476.200	500.000	-4.8
Aroclor-1016-2	5.661	5.566	5.766	500.870	500.000	0.2
Aroclor-1016-3	5.723	5.629	5.829	476.170	500.000	-4.8
Aroclor-1016-4	5.821	5.726	5.926	488.130	500.000	-2.4
Aroclor-1016-5	6.113	6.018	6.218	484.430	500.000	-3.1
Aroclor-1260-1	7.230	7.134	7.334	484.390	500.000	-3.1
Aroclor-1260-2	7.484	7.389	7.589	474.470	500.000	-5.1
Aroclor-1260-3	7.842	7.746	7.946	451.690	500.000	-9.7
Aroclor-1260-4	8.065	7.971	8.171	432.530	500.000	-13.5
Aroclor-1260-5	8.384	8.289	8.489	440.260	500.000	-11.9
Decachlorobiphenyl	10.178	10.086	10.286	45.850	50.000	-8.3
Tetrachloro-m-xylene	4.489	4.393	4.593	48.710	50.000	-2.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2484
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/01/2025 07/01/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073478.D **Time Analyzed:** 23:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.861	4.764	4.964	553.770	500.000	10.8
Aroclor-1016-2	4.879	4.782	4.982	561.550	500.000	12.3
Aroclor-1016-3	5.056	4.958	5.158	576.960	500.000	15.4
Aroclor-1016-4	5.097	5.000	5.200	578.510	500.000	15.7
Aroclor-1016-5	5.311	5.214	5.414	591.760	500.000	18.4
Aroclor-1260-1	6.342	6.244	6.444	550.610	500.000	10.1
Aroclor-1260-2	6.529	6.433	6.633	510.620	500.000	2.1
Aroclor-1260-3	6.682	6.585	6.785	544.940	500.000	9.0
Aroclor-1260-4	7.151	7.054	7.254	545.010	500.000	9.0
Aroclor-1260-5	7.394	7.297	7.497	520.370	500.000	4.1
Decachlorobiphenyl	8.786	8.690	8.890	54.310	50.000	8.6
Tetrachloro-m-xylene	3.783	3.685	3.885	53.770	50.000	7.5