

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070125

Review By	yogesh	Review On	7/1/2025 1:43:36 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:45 AM
SubDirectory	PP070125	HP Acquire Method	HP Processing Method PP070125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP073411.D	01 Jul 2025 13:30	YP\AJ	Ok
2	I.BLK	PP073412.D	01 Jul 2025 13:46	YP\AJ	Ok
3	AR1660ICC1000	PP073413.D	01 Jul 2025 14:04	YP\AJ	Ok
4	AR1660ICC750	PP073414.D	01 Jul 2025 14:21	YP\AJ	Ok
5	AR1660ICC500	PP073415.D	01 Jul 2025 14:37	YP\AJ	Ok
6	AR1660ICC250	PP073416.D	01 Jul 2025 14:54	YP\AJ	Ok
7	AR1660ICC050	PP073417.D	01 Jul 2025 15:10	YP\AJ	Ok,M
8	AR1221ICC500	PP073418.D	01 Jul 2025 15:26	YP\AJ	Ok
9	AR1232ICC500	PP073419.D	01 Jul 2025 15:43	YP\AJ	Ok
10	AR1242ICC1000	PP073420.D	01 Jul 2025 15:59	YP\AJ	Ok
11	AR1242ICC750	PP073421.D	01 Jul 2025 16:16	YP\AJ	Ok
12	AR1242ICC500	PP073422.D	01 Jul 2025 16:33	YP\AJ	Ok
13	AR1242ICC250	PP073423.D	01 Jul 2025 16:49	YP\AJ	Ok
14	AR1242ICC050	PP073424.D	01 Jul 2025 17:05	YP\AJ	Ok,M
15	AR1248ICC1000	PP073425.D	01 Jul 2025 17:22	YP\AJ	Ok
16	AR1248ICC750	PP073426.D	01 Jul 2025 17:39	YP\AJ	Ok
17	AR1248ICC500	PP073427.D	01 Jul 2025 17:55	YP\AJ	Ok
18	AR1248ICC250	PP073428.D	01 Jul 2025 18:12	YP\AJ	Ok
19	AR1248ICC050	PP073429.D	01 Jul 2025 18:28	YP\AJ	Ok,M
20	AR1254ICC1000	PP073430.D	01 Jul 2025 18:45	YP\AJ	Ok
21	AR1254ICC750	PP073431.D	01 Jul 2025 19:01	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070125

Review By	yogesh	Review On	7/1/2025 1:43:36 PM
Supervise By	mohammad	Supervise On	7/2/2025 7:44:45 AM
SubDirectory	PP070125	HP Acquire Method	HP Processing Method PP070125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP073432.D	01 Jul 2025 19:18	YP\AJ	Ok
23	AR1254ICC250	PP073433.D	01 Jul 2025 19:35	YP\AJ	Ok
24	AR1254ICC050	PP073434.D	01 Jul 2025 19:51	YP\AJ	Ok,M
25	AR1262ICC500	PP073435.D	01 Jul 2025 20:08	YP\AJ	Ok
26	AR1268ICC1000	PP073436.D	01 Jul 2025 20:24	YP\AJ	Ok
27	AR1268ICC750	PP073437.D	01 Jul 2025 20:41	YP\AJ	Ok
28	AR1268ICC500	PP073438.D	01 Jul 2025 20:57	YP\AJ	Ok
29	AR1268ICC250	PP073439.D	01 Jul 2025 21:14	YP\AJ	Ok,M
30	AR1268ICC050	PP073440.D	01 Jul 2025 21:30	YP\AJ	Ok,M
31	PP070125ICV500	PP073441.D	01 Jul 2025 21:47	YP\AJ	Ok
32	AR1242ICV500	PP073442.D	01 Jul 2025 22:20	YP\AJ	Ok
33	AR1248ICV500	PP073443.D	01 Jul 2025 22:53	YP\AJ	Ok
34	AR1254ICV500	PP073444.D	01 Jul 2025 23:26	YP\AJ	Ok
35	AR1268ICV500	PP073445.D	02 Jul 2025 00:00	YP\AJ	Ok
36	DDT ANALOGUE	PP073446.D	02 Jul 2025 07:34	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070225

Review By	yogesh	Review On	7/2/2025 10:58:10 AM
Supervise By	mohammad	Supervise On	7/4/2025 4:31:38 AM
SubDirectory	PP070225	HP Acquire Method	HP Processing Method PP070125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP073447.D	02 Jul 2025 08:57	YP\AJ	Ok
2	AR1660CCC500	PP073448.D	02 Jul 2025 10:07	YP\AJ	Ok
3	AR1242CCC500	PP073449.D	02 Jul 2025 10:27	YP\AJ	Ok
4	AR1248CCC500	PP073450.D	02 Jul 2025 10:44	YP\AJ	Ok
5	AR1254CCC500	PP073451.D	02 Jul 2025 11:00	YP\AJ	Ok
6	I.BLK	PP073452.D	02 Jul 2025 11:17	YP\AJ	Ok
7	DDT ANALOGUE	PP073453.D	02 Jul 2025 11:34	YP\AJ	Ok
8	PB168692BL	PP073454.D	02 Jul 2025 15:14	YP\AJ	Ok
9	PB168692BS	PP073455.D	02 Jul 2025 15:30	YP\AJ	Ok
10	Q2472-01	PP073456.D	02 Jul 2025 15:46	YP\AJ	Ok
11	Q2473-01	PP073457.D	02 Jul 2025 16:03	YP\AJ	Ok,M
12	Q2473-02	PP073458.D	02 Jul 2025 16:19	YP\AJ	Ok,M
13	Q2473-03	PP073459.D	02 Jul 2025 16:36	YP\AJ	Ok,M
14	Q2473-04	PP073460.D	02 Jul 2025 16:53	YP\AJ	Ok,M
15	Q2473-04MS	PP073461.D	02 Jul 2025 17:09	YP\AJ	Ok,M
16	Q2473-04MSD	PP073462.D	02 Jul 2025 17:25	YP\AJ	Ok,M
17	AR1660CCC500	PP073463.D	02 Jul 2025 18:47	YP\AJ	Ok,M
18	AR1242CCC500	PP073464.D	02 Jul 2025 19:04	YP\AJ	Ok
19	AR1248CCC500	PP073465.D	02 Jul 2025 19:20	YP\AJ	Ok
20	AR1254CCC500	PP073466.D	02 Jul 2025 19:36	YP\AJ	Ok
21	I.BLK	PP073467.D	02 Jul 2025 19:53	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP070225

Review By	yogesh	Review On	7/2/2025 10:58:10 AM
Supervise By	mohammad	Supervise On	7/4/2025 4:31:38 AM
SubDirectory	PP070225	HP Acquire Method	HP Processing Method PP070125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q2484-01	PP073468.D	02 Jul 2025 20:09	YP\AJ	Ok,M
23	Q2484-02	PP073469.D	02 Jul 2025 20:26	YP\AJ	Ok,M
24	Q2484-03	PP073470.D	02 Jul 2025 20:42	YP\AJ	Ok
25	Q2484-04	PP073471.D	02 Jul 2025 20:58	YP\AJ	Ok
26	Q2484-05	PP073472.D	02 Jul 2025 21:15	YP\AJ	Ok
27	Q2484-06	PP073473.D	02 Jul 2025 21:31	YP\AJ	Ok
28	Q2458-10	PP073474.D	02 Jul 2025 21:47	YP\AJ	Ok
29	Q2477-01	PP073475.D	02 Jul 2025 22:04	YP\AJ	Ok,M
30	Q2481-01	PP073476.D	02 Jul 2025 22:20	YP\AJ	Ok,M
31	Q2481-04	PP073477.D	02 Jul 2025 22:36	YP\AJ	Ok,M
32	AR1660CCC500	PP073478.D	02 Jul 2025 23:58	YP\AJ	Ok,M
33	AR1242CCC500	PP073479.D	03 Jul 2025 00:15	YP\AJ	Not Ok
34	AR1248CCC500	PP073480.D	03 Jul 2025 00:31	YP\AJ	Ok
35	AR1254CCC500	PP073481.D	03 Jul 2025 00:47	YP\AJ	Not Ok
36	I.BLK	PP073482.D	03 Jul 2025 01:04	YP\AJ	Ok

M : Manual Integration