

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bf070825

SequenceID : <u>bf070825</u>	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	<u>✓</u>
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	<u>✓</u>
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	<u>✓</u>
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	<u>✓</u>
5. GC/MS Calibration Met:	_____	_____	<u>✓</u>
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	<u>✓</u>
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. Compounds #54,56,70,77 are biased high in the CCC (BF143025.D), if any samples are found with hit of these compounds they will be re-analyzed with a properly passing CCC.	_____	_____	<u>✓</u>
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	<u>✓</u>	_____
a. B/N Fraction			
d. Acid Fraction			

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

a. B/N Fraction

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

a. B/N Fraction PB168716BSD : RPD failed due to result difference between BS and BSD.
Q2514-08MS/MSD recovery failed due to matrix interference. No corrective action is required.

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

ADDITIONAL COMMENTS:

Butylbenzylphthalate is marginally biased high in the PB168716BS. The data will be used for hardcopies.

Pratik

Analyst

07/09/2025

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bf070825

SequenceID : bf070825 NA NO YES