

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF061925

Review By	Rahul	Review On	6/23/2025 3:55:52 PM
Supervise By	Jagrut	Supervise On	6/24/2025 3:17:49 PM
SubDirectory	BF061925	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12670,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142787.D	19 Jun 2025 16:37		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF142788.D	19 Jun 2025 17:07		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF142789.D	19 Jun 2025 17:38		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF142790.D	19 Jun 2025 18:08		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF142791.D	19 Jun 2025 18:39		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF142792.D	19 Jun 2025 19:09		RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF142793.D	19 Jun 2025 19:40		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF142794.D	19 Jun 2025 20:10		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF142795.D	19 Jun 2025 20:40	Compound #09 removed from 80 PPM.	RC/JU	Ok
10	SSTDICV040	ICVBF061925	BF142796.D	19 Jun 2025 21:10		RC/JU	Ok
11	PB168536BL	PB168536BL	BF142797.D	19 Jun 2025 22:10		RC/JU	Ok
12	PB168536BS	PB168536BS	BF142798.D	19 Jun 2025 22:40		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070225

Review By	Rahul	Review On	7/3/2025 9:14:42 AM
Supervise By	Jagrut	Supervise On	7/3/2025 11:11:28 AM
SubDirectory	BF070225	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12672,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142963.D	02 Jul 2025 11:11		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142964.D	02 Jul 2025 11:41		RC/JU	Ok
3	PB168652TB	PB168652TB	BF142965.D	02 Jul 2025 12:11		RC/JU	Ok
4	Q2452-04	TP-5	BF142966.D	02 Jul 2025 12:46		RC/JU	Ok
5	Q2452-04MS	TP-5MS	BF142967.D	02 Jul 2025 13:16		RC/JU	Ok,M
6	Q2452-04MSD	TP-5MSD	BF142968.D	02 Jul 2025 13:47		RC/JU	Ok,M
7	Q2458-04	TP-67	BF142969.D	02 Jul 2025 14:17		RC/JU	Ok
8	Q2458-04MS	TP-67MS	BF142970.D	02 Jul 2025 14:47		RC/JU	Ok,M
9	Q2458-04MSD	TP-67MSD	BF142971.D	02 Jul 2025 15:17		RC/JU	Ok,M
10	Q2452-08	EP-2	BF142972.D	02 Jul 2025 15:47		RC/JU	Ok
11	Q2459-04	TP-1	BF142973.D	02 Jul 2025 16:18		RC/JU	Ok
12	Q2469-04	WC-1	BF142974.D	02 Jul 2025 16:49		RC/JU	Ok
13	Q2458-09	TP-59	BF142975.D	02 Jul 2025 17:19		RC/JU	Ok
14	Q2459-01	TP-1	BF142976.D	02 Jul 2025 17:49		RC/JU	Ok
15	Q2458-05	TP-66	BF142977.D	02 Jul 2025 18:20		RC/JU	Ok,M
16	Q2458-01	TP-76	BF142978.D	02 Jul 2025 18:50		RC/JU	Ok,M
17	Q2458-03	TP-68	BF142979.D	02 Jul 2025 19:21		RC/JU	Ok,M
18	Q2450-01DL	PL-2-062725DL	BF142980.D	02 Jul 2025 19:52		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070225

Review By	Rahul	Review On	7/3/2025 9:14:42 AM		
Supervise By	Jagrut	Supervise On	7/3/2025 11:11:28 AM		
SubDirectory	BF070225	HP Acquire Method	BNA_F	HP Processing Method	bf061925
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S12672,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2475-01	SOIL-PILE	BF142981.D	02 Jul 2025 20:22		RC/JU	Ok,M
20	Q2475-01MS	SOIL-PILEMS	BF142982.D	02 Jul 2025 20:52		RC/JU	Ok,M
21	Q2475-01MSD	SOIL-PILEMSD	BF142983.D	02 Jul 2025 21:22		RC/JU	Ok,M
22	Q2478-01	WC-1	BF142984.D	02 Jul 2025 21:52		RC/JU	Ok
23	Q2458-02	TP-55	BF142985.D	02 Jul 2025 22:22		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070325

Review By	Rahul	Review On	7/7/2025 11:33:04 AM
Supervise By	Jagrut	Supervise On	7/7/2025 12:17:49 PM
SubDirectory	BF070325	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12672,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142986.D	03 Jul 2025 09:25		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142987.D	03 Jul 2025 09:55		RC/JU	Ok,M
3	PB168694BL	PB168694BL	BF142988.D	03 Jul 2025 10:26		RC/JU	Ok
4	PB168694BS	PB168694BS	BF142989.D	03 Jul 2025 10:55		RC/JU	Ok,M
5	PB168674BL	PB168674BL	BF142990.D	03 Jul 2025 11:25		RC/JU	Ok
6	PB168674BS	PB168674BS	BF142991.D	03 Jul 2025 11:55		RC/JU	Ok,M
7	PB168685BL	PB168685BL	BF142992.D	03 Jul 2025 12:25		RC/JU	Ok
8	PB168685BS	PB168685BS	BF142993.D	03 Jul 2025 12:54		RC/JU	Ok,M
9	Q2472-01	PN#1	BF142994.D	03 Jul 2025 13:55		RC/JU	Ok
10	Q2484-02	TP-57	BF142995.D	03 Jul 2025 14:25		RC/JU	Ok,M
11	Q2484-06	TP-104	BF142996.D	03 Jul 2025 14:56		RC/JU	Ok
12	Q2484-05	TP-106	BF142997.D	03 Jul 2025 15:27		RC/JU	Ok
13	Q2458-10	FB-06272025	BF142998.D	03 Jul 2025 15:58		RC/JU	Ok
14	Q2463-01	TW-WTS-11	BF142999.D	03 Jul 2025 16:29		RC/JU	Ok
15	Q2484-04	TP-107	BF143000.D	03 Jul 2025 16:59		RC/JU	Ok
16	Q2473-04	PIT#4	BF143001.D	03 Jul 2025 17:30	Parameter list not updated on login page	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070725

Review By	Rahul	Review On	7/8/2025 10:39:40 AM
Supervise By	Jagrut	Supervise On	7/8/2025 11:33:00 AM
SubDirectory	BF070725	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12672,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143002.D	07 Jul 2025 11:12		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143003.D	07 Jul 2025 12:12		RC/JU	Ok,M
3	PB168722BL	PB168722BL	BF143004.D	07 Jul 2025 12:42		RC/JU	Ok
4	Q2493-01	WC-11	BF143005.D	07 Jul 2025 13:16		RC/JU	Ok
5	Q2493-01MS	WC-11MS	BF143006.D	07 Jul 2025 13:47		RC/JU	Ok,M
6	Q2493-01MSD	WC-11MSD	BF143007.D	07 Jul 2025 14:17		RC/JU	Ok,M
7	Q2487-12	G3(6-12)	BF143008.D	07 Jul 2025 14:47		RC/JU	Ok
8	Q2487-10	G4(6-12)	BF143009.D	07 Jul 2025 15:17		RC/JU	Ok,M
9	Q2487-14	G2(6-12)	BF143010.D	07 Jul 2025 15:48		RC/JU	Ok,M
10	Q2491-01	EO-1-070225	BF143011.D	07 Jul 2025 16:18		RC/JU	Ok,M
11	Q2478-04	WC-1	BF143012.D	07 Jul 2025 16:49		RC/JU	Ok
12	Q2478-04MS	WC-1MS	BF143013.D	07 Jul 2025 17:19		RC/JU	Ok,M
13	Q2478-04MSD	WC-1MSD	BF143014.D	07 Jul 2025 17:50		RC/JU	Ok,M
14	Q2514-02	TP-93	BF143015.D	07 Jul 2025 18:21		RC/JU	Ok
15	Q2513-03	HR-3-070325	BF143016.D	07 Jul 2025 18:51		RC/JU	Ok,M
16	Q2507-01	SU-04-7.3-2025	BF143017.D	07 Jul 2025 19:21		RC/JU	Ok,M
17	Q2484-01	TP-58	BF143018.D	07 Jul 2025 19:51	Need 5X Dilution	RC/JU	Dilution

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070725

Review By	Rahul	Review On	7/8/2025 10:39:40 AM		
Supervise By	Jagrut	Supervise On	7/8/2025 11:33:00 AM		
SubDirectory	BF070725	HP Acquire Method	BNA_F	HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12672,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q2473-03	PIT#3	BF143019.D	07 Jul 2025 20:21	Parameter list not updated on login page	RC/JU	Ok,M
19	Q2473-01	PIT#1	BF143020.D	07 Jul 2025 20:51	need DL 10X	RC/JU	Ok,M
20	Q2484-03	TP-64	BF143021.D	07 Jul 2025 21:22		RC/JU	Ok,M
21	Q2473-02	PIT#2	BF143022.D	07 Jul 2025 21:52	Parameter list not updated on login page	RC/JU	Ok,M
22	Q2513-01	HR-2-070325	BF143023.D	07 Jul 2025 22:22		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070825

Review By	Rahul	Review On	7/9/2025 9:47:19 AM
Supervise By	Jagrut	Supervise On	7/9/2025 2:16:38 PM
SubDirectory	BF070825	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143024.D	08 Jul 2025 10:24		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143025.D	08 Jul 2025 10:54		RC/JU	Ok,M
3	PB168716BL	PB168716BL	BF143026.D	08 Jul 2025 11:24		RC/JU	Ok
4	PB168716BS	PB168716BS	BF143027.D	08 Jul 2025 11:54		RC/JU	Ok,M
5	PB168716BSD	PB168716BSD	BF143028.D	08 Jul 2025 12:25		RC/JU	Ok,M
6	PB168722BS	PB168722BS	BF143029.D	08 Jul 2025 12:55		RC/JU	Ok,M
7	Q2514-07	TP-36	BF143030.D	08 Jul 2025 13:31		RC/JU	Ok
8	Q2514-08	TP-78	BF143031.D	08 Jul 2025 14:02		RC/JU	Ok
9	Q2514-08MS	TP-78MS	BF143032.D	08 Jul 2025 14:32		RC/JU	Ok,M
10	Q2514-08MSD	TP-78MSD	BF143033.D	08 Jul 2025 15:03		RC/JU	Ok,M
11	Q2515-01	wc-1	BF143034.D	08 Jul 2025 15:34		RC/JU	Ok
12	Q2514-09	TP-81	BF143035.D	08 Jul 2025 16:05		RC/JU	Ok
13	Q2514-03	TP-94	BF143036.D	08 Jul 2025 16:36		RC/JU	Ok
14	Q2514-01	TP-92	BF143037.D	08 Jul 2025 17:07		RC/JU	Ok
15	Q2514-05	TP-97	BF143038.D	08 Jul 2025 17:37		RC/JU	Ok,M
16	Q2489-01	G4(0-6)	BF143039.D	08 Jul 2025 18:08		RC/JU	Ok
17	Q2489-02	G4(6-12)	BF143040.D	08 Jul 2025 18:39		RC/JU	Ok
18	Q2493-04	WC-11	BF143041.D	08 Jul 2025 19:10		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070825

Review By	Rahul	Review On	7/9/2025 9:47:19 AM		
Supervise By	Jagrut	Supervise On	7/9/2025 2:16:38 PM		
SubDirectory	BF070825	HP Acquire Method	BNA_F	HP Processing Method	bf061925
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S12673,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2484-01DL	TP-58DL	BF143042.D	08 Jul 2025 19:41		RC/JU	Ok,M
20	Q2473-01DL	PIT#1DL	BF143043.D	08 Jul 2025 20:12	Parameter list not updated on login page	RC/JU	Ok,M
21	Q2514-06	TP-103	BF143044.D	08 Jul 2025 20:42		RC/JU	Ok
22	Q2514-10	TP-90	BF143045.D	08 Jul 2025 21:12		RC/JU	Ok,M
23	Q2514-04	TP-96	BF143046.D	08 Jul 2025 21:42		RC/JU	Ok,M

M : Manual Integration