

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2484

Client ID : SSTDCCC040

Date Analyzed: 07/02/2025

Lab File ID: BF142964.D

Time Analyzed: 11:41

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	58835	6.875	223613	8.16	116653	9.92
UPPER LIMIT	117670	7.375	447226	8.657	233306	10.416
LOWER LIMIT	29417.5	6.375	111807	7.657	58326.5	9.416
EPA SAMPLE NO.						
01 SOIL-PILEMS	59177	6.88	222831	8.16	115354	9.92
02 SOIL-PILEMSD	59335	6.88	228314	8.16	114108	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.