

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2484

Client ID : SSTDCCC040

Date Analyzed: 07/07/2025

Lab File ID: BF143003.D

Time Analyzed: 12:12

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	66299	6.875	250576	8.16	128386	9.92
UPPER LIMIT	132598	7.375	501152	8.663	256772	10.422
LOWER LIMIT	33149.5	6.375	125288	7.663	64193	9.422
EPA SAMPLE NO.						
01 TP-58	47809	6.88	165637	8.16	87479	9.92
02 TP-64	45344	6.88	158652	8.16	77572	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.