

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF142986.D
Instrument ID: BNA_F

Contract: CAMP02
SDG NO.: Q2484
DFTPP Injection Date: 07/03/2025
DFTPP Injection Time: 09:25

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	30.1
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	28.4
70	Less than 2.0% of mass 69	0.1 (0.3) 1
127	10.0 - 80.0% of mass 198	39
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.4
275	10.0 - 60.0% of mass 198	24.2
365	Greater than 1% of mass 198	3.2
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.8 (18.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142987.D	07/03/2025	09:55
PB168694BL	PB168694BL	BF142988.D	07/03/2025	10:26
PB168694BS	PB168694BS	BF142989.D	07/03/2025	10:55
TP-57	Q2484-02	BF142995.D	07/03/2025	14:25
TP-104	Q2484-06	BF142996.D	07/03/2025	14:56
TP-106	Q2484-05	BF142997.D	07/03/2025	15:27
TP-107	Q2484-04	BF143000.D	07/03/2025	16:59