

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CAMP02
 Lab Code: ACE SDG NO.: Q2484
 Lab File ID: VY022946.D Date Analyzed: 07/07/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:08
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	469181	7.71	754142	8.62	648632	11.41
UPPER LIMIT	938362	8.207	1508280	9.116	1297260	11.914
LOWER LIMIT	234591	7.207	377071	8.116	324316	10.914
EPA SAMPLE NO.						
TP-58	303940	7.71	573765	8.62	578554	11.41
TP-57	301710	7.71	565210	8.62	552333	11.41
TP-64	287261	7.71	532666	8.62	486729	11.41
TP-107	324819	7.71	601393	8.62	567117	11.41
TP-104	266886	7.71	509158	8.62	510222	11.41
VY0707SBL01	359794	7.71	659995	8.62	632698	11.41
VY0707SBS01	438661	7.71	730129	8.61	612594	11.41
VY0707SBSD01	433247	7.71	718134	8.62	614643	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.