

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q2485

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB136364

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.097	0.099	98	85-115	07/02/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	07/02/2025
Sample ID: Cyanide	CCV2	mg/L	0.26	0.25	104	90-110	07/02/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q2485

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB136464

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.02	7	100	90-110	07/01/2025
Sample ID:	CCV1						
Field pH		pH	7.01	7	100	90-110	07/01/2025
Sample ID:	CCV2						
Field pH		pH	7.01	7	100	90-110	07/01/2025