

Initial and Continuing Calibration Blank Summary

Client: METEM A GE POWER Business

SDG No.: Q2485

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB136364

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	07/02/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	07/02/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	07/02/2025