

Preparation Blank Summary

Client: METEM A GE POWER Business

SDG No.: Q2485

Project: Parsippany Wastewater Quarterly 2025

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: Cyanide	PB168703BL mg/L	< 0.0025	0.0025	U	0.0012	0.005	07/02/2025