

Laboratory Control Sample Summary

Client: METEM A GE POWER Business

SDG No.: Q2485

Project: Parsippany Wastewater Quarterly 2025

Run No.: LB136364

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168703BS							
Cyanide	mg/L	0.1	0.098		98	1	85-115	07/02/2025