

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136487

Review By	Janvi	Review On	7/16/2025 1:08:48 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212
ICV Standard	MP86213
CCV Standard	MP86216
ICSA Standard	MP86214,MP86215
CRI Standard	MP86212
LCS Standard	
Chk Standard	MP86217,MP86218

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/15/25 12:32		Jaswal	OK
2	S1	S1	CAL2	07/15/25 12:37		Jaswal	OK
3	S2	S2	CAL3	07/15/25 12:41		Jaswal	OK
4	S3	S3	CAL4	07/15/25 12:45		Jaswal	OK
5	S4	S4	CAL5	07/15/25 12:49		Jaswal	OK
6	S5	S5	CAL6	07/15/25 12:53		Jaswal	OK
7	ICV01	ICV01	ICV	07/15/25 13:11		Jaswal	OK
8	LLICV01	LLICV01	LLICV	07/15/25 13:16		Jaswal	OK
9	ICB01	ICB01	ICB	07/15/25 13:21		Jaswal	OK
10	CRI01	CRI01	CRDL	07/15/25 13:25		Jaswal	OK
11	ICSA01	ICSA01	ICSA	07/15/25 13:29		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	07/15/25 13:43		Jaswal	OK
13	ICSADL	ICSADL	ICSA	07/15/25 13:47		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	07/15/25 13:51		Jaswal	OK
15	CCV01	CCV01	CCV	07/15/25 13:55		Jaswal	OK
16	CCB01	CCB01	CCB	07/15/25 14:00		Jaswal	OK
17	Q2485-01	WASTEWATER Q3	SAM	07/15/25 14:04		Jaswal	OK
18	Q2485-01DUP	WASTEWATER Q3DUP	DUP	07/15/25 14:09		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136487

Review By	Janvi	Review On	7/16/2025 1:08:48 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212
ICV Standard	MP86213
CCV Standard	MP86216
ICSA Standard	MP86214,MP86215
CRI Standard	MP86212
LCS Standard	
Chk Standard	MP86217,MP86218

19	Q2485-01L	WASTEWATER Q3L	SD	07/15/25 14:13		Jaswal	OK
20	Q2485-01MS	WASTEWATER Q3MS	MS	07/15/25 14:18	MS/MSD fail for more than 50% parameters	Jaswal	OK
21	Q2485-01MSD	WASTEWATER Q3MSD	MSD	07/15/25 14:22	MS/MSD fail for more than 50% parameters	Jaswal	OK
22	Q2485-01A	WASTEWATER Q3A	PS	07/15/25 14:26	0.1ml of M6013,M6004 of 10ml sample.	Jaswal	OK
23	PB168715BS	PB168715BS	LCS	07/15/25 14:56		Jaswal	OK
24	PB168715BL	PB168715BL	MB	07/15/25 15:27		Jaswal	OK
25	Q2593-01	001 WILLETS PT BLV	SAM	07/15/25 15:31		Jaswal	OK
26	Q2593-02	002 35th AVE (MAY)	SAM	07/15/25 15:35		Jaswal	OK
27	CCV02	CCV02	CCV	07/15/25 15:46	CCV fail for Na	Jaswal	OK
28	CCB02	CCB02	CCB	07/15/25 15:51	CCB fail for Na	Jaswal	OK