



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: METEM A GE POWER Business

SDG No.: Q2485

Contract: METE01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.82	6.0	97	65 - 135	P	07/15/2025	13:25	LB136487
	Chromium	9.88	10.0	99	65 - 135	P	07/15/2025	13:25	LB136487
	Copper	21.4	20.0	107	65 - 135	P	07/15/2025	13:25	LB136487
	Lead	11.7	12.0	97	65 - 135	P	07/15/2025	13:25	LB136487
	Nickel	40.2	40.0	101	65 - 135	P	07/15/2025	13:25	LB136487
	Silver	10.3	10.0	102	65 - 135	P	07/15/2025	13:25	LB136487
	Zinc	39.8	40.0	100	65 - 135	P	07/15/2025	13:25	LB136487