



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## Metals

- 3a -

### INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q2485

Contract: METE01

Lab Code: ACE

| Sample ID | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| ICB01     | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 07/15/2025       | 13:21            | LB136487      |
|           | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 07/15/2025       | 13:21            | LB136487      |
|           | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/15/2025       | 13:21            | LB136487      |
|           | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 07/15/2025       | 13:21            | LB136487      |
|           | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/15/2025       | 13:21            | LB136487      |
|           | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 07/15/2025       | 13:21            | LB136487      |
|           | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/15/2025       | 13:21            | LB136487      |

**Metals**

- 3a -

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** METEM A GE POWER Business

**SDG No.:** Q2485

**Contract:** METE01

**Lab Code:** ACE

| Sample ID    | Analyte  | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|--------------|----------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
| <b>CCB01</b> | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 07/15/2025       | 14:00            | LB136487      |
|              | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 07/15/2025       | 14:00            | LB136487      |
|              | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/15/2025       | 14:00            | LB136487      |
|              | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 07/15/2025       | 14:00            | LB136487      |
|              | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/15/2025       | 14:00            | LB136487      |
|              | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 07/15/2025       | 14:00            | LB136487      |
|              | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/15/2025       | 14:00            | LB136487      |
| <b>CCB02</b> | Cadmium  | 0.50           | +/-3                | U            | 6.00 | P | 07/15/2025       | 15:51            | LB136487      |
|              | Chromium | 2.12           | +/-5                | U            | 10.0 | P | 07/15/2025       | 15:51            | LB136487      |
|              | Copper   | 4.60           | +/-10               | U            | 20.0 | P | 07/15/2025       | 15:51            | LB136487      |
|              | Lead     | 2.30           | +/-6                | U            | 12.0 | P | 07/15/2025       | 15:51            | LB136487      |
|              | Nickel   | 3.06           | +/-20               | U            | 40.0 | P | 07/15/2025       | 15:51            | LB136487      |
|              | Silver   | 1.62           | +/-5                | U            | 10.0 | P | 07/15/2025       | 15:51            | LB136487      |
|              | Zinc     | 16.7           | +/-20               | U            | 40.0 | P | 07/15/2025       | 15:51            | LB136487      |

**Metals**

**- 3a -**

**INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY**

**Client:** METEM A GE POWER Business

**SDG No.:** Q2485

**Contract:** METE01

**Lab Code:** ACE

| Sample ID | Analyte | Result<br>ug/L | Acceptance<br>Limit | Conc<br>Qual | CRQL | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|
|-----------|---------|----------------|---------------------|--------------|------|---|------------------|------------------|---------------|