

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q2485

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168715BL	WATER			Batch Number:	PB168715		Prep Date:	07/03/2025	
	Cadmium	0.31	<1.5	U	3.00	P	07/15/2025	15:27	LB136487
	Chromium	0.53	<2.5	U	5.00	P	07/15/2025	15:27	LB136487
	Copper	1.89	<5	U	10.0	P	07/15/2025	15:27	LB136487
	Lead	1.21	<3	U	6.00	P	07/15/2025	15:27	LB136487
	Nickel	1.90	<10	U	20.0	P	07/15/2025	15:27	LB136487
	Silver	0.81	<2.5	U	5.00	P	07/15/2025	15:27	LB136487
	Zinc	2.00	<10	U	20.0	P	07/15/2025	15:27	LB136487