

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>METEM A GE POWER Business</u>	SDG No.: <u>Q2485</u>
Contract: <u>METE01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-0.039	1.0	4	-5	7	07/15/2025	13:29	LB136487
	Chromium	56.8	52.0	109	42	62	07/15/2025	13:29	LB136487
	Copper	2.98	2.0	149	-18	22	07/15/2025	13:29	LB136487
	Lead	-6.97			-12	12	07/15/2025	13:29	LB136487
	Nickel	18.9	2.0	945	-38	42	07/15/2025	13:29	LB136487
	Silver	1.03			-10	10	07/15/2025	13:29	LB136487
	Zinc	4.07			-40	40	07/15/2025	13:29	LB136487
ICSAB01	Cadmium	977	972	100	826	1120	07/15/2025	13:43	LB136487
	Chromium	518	542	96	460	624	07/15/2025	13:43	LB136487
	Copper	481	511	94	434	588	07/15/2025	13:43	LB136487
	Lead	38.5	49.0	79	37	61	07/15/2025	13:43	LB136487
	Nickel	982	954	103	810	1100	07/15/2025	13:43	LB136487
	Silver	215	201	107	170	232	07/15/2025	13:43	LB136487
	Zinc	1060	952	111	809	1095	07/15/2025	13:43	LB136487
ICSA	Cadmium	-1.18	1.0	118	-5	7	07/15/2025	13:47	LB136487
	Chromium	64.4	52.0	124	42	62	07/15/2025	13:47	LB136487
	Copper	17.3	2.0	865	-18	22	07/15/2025	13:47	LB136487
	Lead	-18.7			-12	12	07/15/2025	13:47	LB136487
	Nickel	25.5	2.0	1275	-38	42	07/15/2025	13:47	LB136487
	Silver	-9.72			-10	10	07/15/2025	13:47	LB136487
	Zinc	0.24			-40	40	07/15/2025	13:47	LB136487
ICSAB	Cadmium	953	972	98	826	1120	07/15/2025	13:51	LB136487
	Chromium	533	542	98	460	624	07/15/2025	13:51	LB136487
	Copper	509	511	100	434	588	07/15/2025	13:51	LB136487
	Lead	19.0	49.0	39	37	61	07/15/2025	13:51	LB136487
	Nickel	972	954	102	810	1100	07/15/2025	13:51	LB136487
	Silver	193	201	96	170	232	07/15/2025	13:51	LB136487
	Zinc	1030	952	108	809	1095	07/15/2025	13:51	LB136487