

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q2485

Contract: METE01

Lab Code: ACE

Matrix: Water

Level: LOW

Client ID: WASTEWATER Q3A

Sample ID: Q2485-01 **Spiked ID:** Q2485-01A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	85 - 115	70.3		3.00	U	100	70	N	P
Chromium	ug/L	85 - 115	118		14.8		200	52	N	P
Copper	ug/L	85 - 115	82.5		10.0	U	150	55	N	P
Lead	ug/L	85 - 115	343		6.00	U	500	69	N	P
Nickel	ug/L	85 - 115	351		169		250	73	N	P
Silver	ug/L	85 - 115	93.8		1.27	J	37.5	247	N	P
Zinc	ug/L	85 - 115	256		112		100	144	N	P