

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: METEM A GE POWER Business **Level:** LOW **SDG No.:** Q2485
Contract: METE01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2485-01 **Client ID:** WASTEWATER Q3DUP
Percent Solids for Sample: NA **Duplicate ID** Q2485-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	3.00	U	3.00	U			P
Chromium	ug/L	20	14.8		15.2		3		P
Copper	ug/L	20	10.0	U	10.0	U			P
Lead	ug/L	20	6.00	U	6.00	U			P
Nickel	ug/L	20	169		173		2		P
Silver	ug/L	20	1.27	J	1.03	J	21		P
Zinc	ug/L	20	112		44.8		86	*	P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: METEM A GE POWER Business **Level:** LOW **SDG No.:** Q2485
Contract: METE01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2485-01MS **Client ID:** WASTEWATER Q3MSD
Percent Solids for Sample: NA **Duplicate ID** Q2485-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	72.6		71.5		2		P
Chromium	ug/L	20	124		124		0		P
Copper	ug/L	20	86.5		86.1		0		P
Lead	ug/L	20	352		347		1		P
Nickel	ug/L	20	369		347		6		P
Silver	ug/L	20	95.9		97.6		2		P
Zinc	ug/L	20	276		271		2		P