

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q2485

Contract: METE01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168715BS							
Cadmium	ug/L	100	96.0		96	85 - 115	P
Chromium	ug/L	200	201		100	85 - 115	P
Copper	ug/L	150	153		102	85 - 115	P
Lead	ug/L	500	481		96	85 - 115	P
Nickel	ug/L	250	247		99	85 - 115	P
Silver	ug/L	37.5	37.4		100	85 - 115	P
Zinc	ug/L	100	112		112	85 - 115	P