

Continuing Calibration Report for SequenceID : FC070825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069388.D

Aliphatic C9-C12	8616465.000	60.000	3.303	6.602	143607.750	146679.627	2.094
Aliphatic C12-C16	5935838.000	40.000	6.603	10.005	148395.950	159210.532	6.793
Aliphatic C16-C21	8346556.000	60.000	10.006	13.375	139109.267	154434.323	9.923
Aliphatic C21-C28	9796765.000	80.000	13.375	17.040	122459.563	133966.368	8.589
Aliphatic C28-C40	11269119.000	120.000	17.041	22.023	93909.325	92124.417	-1.937
Aliphatic EPH	43964743.000	360.000	3.303	22.023	122124.286	128354.160	4.854