

Continuing Calibration Report for SequenceID : FC070825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069397.D

Aliphatic C9-C12	8595615.000	60.000	3.303	6.602	143260.250	146679.627	2.331
Aliphatic C12-C16	5921848.000	40.000	6.603	10.005	148046.200	159210.532	7.012
Aliphatic C16-C21	8372066.000	60.000	10.006	13.375	139534.433	154434.323	9.648
Aliphatic C21-C28	9935211.000	80.000	13.375	17.040	124190.138	133966.368	7.298
Aliphatic C28-C40	11659559.000	120.000	17.041	22.023	97162.992	92124.417	-5.469
Aliphatic EPH	44484299.000	360.000	3.303	22.023	123567.497	128354.160	3.729