

Continuing Calibration Report for SequenceID : FE070825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE054740.D

Aliphatic C9-C12	8494556.000	60.000	3.323	6.956	141575.933	136055.079	-4.058
Aliphatic C12-C16	5894995.000	40.000	6.957	10.409	147374.875	140568.128	-4.842
Aliphatic C16-C21	8539672.000	60.000	10.410	13.786	142327.867	144393.643	1.431
Aliphatic C21-C28	12069329.000	80.000	13.787	17.458	150866.613	144651.717	-4.296
Aliphatic C28-C40	16504903.000	120.000	17.459	22.482	137540.858	138647.492	0.798
Aliphatic EPH	51503455.000	360.000	3.323	22.482	143065.153	140720.791	-1.666