

Continuing Calibration Report for SequenceID : FE070825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE054750.D

Aliphatic C9-C12	8433339.000	60.000	3.323	6.956	140555.650	136055.079	-3.308
Aliphatic C12-C16	5898098.000	40.000	6.957	10.409	147452.450	140568.128	-4.897
Aliphatic C16-C21	8599688.000	60.000	10.410	13.786	143328.133	144393.643	0.738
Aliphatic C21-C28	12187137.000	80.000	13.787	17.458	152339.213	144651.717	-5.314
Aliphatic C28-C40	17119618.000	120.000	17.459	22.482	142663.483	138647.492	-2.897
Aliphatic EPH	52237880.000	360.000	3.323	22.482	145105.222	140720.791	-3.116